

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE110617\
 Data File : BE094752.D
 Acq On : 6 Nov 2017 14:31
 Operator : SJ/JU
 Sample : I6150-06
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 B-3(2-3)-102517

Manual Integrations
 APPROVED

Sohil
 11/7/2017 5:41:32 PM

Quant Time: Nov 07 00:41:06 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE102517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 07 00:12:32 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	1076	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.69	136	6771	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.51	164	4227	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	8049	0.40	ng/ul	-0.02
16) Chrysene-d12	21.43	240	7980	0.40	ng/ul	0.00
20) Perylene-d12	23.97	264	8120m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.27	152	2713	0.25	ng/ul	-0.02
14) Fluoranthene-d10	19.28	212	6241	0.26	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.73	128	254200	15.355	ng/ul#	93
5) 2-Methylnaphthalene	12.34	142	32078	2.770	ng/ul	99
7) Acenaphthylene	14.24	152	942	0.049	ng/ul#	53
8) Acenaphthene	14.57	153	7129	0.519	ng/ul	97
9) Fluorene	15.57	166	11116	0.720	ng/ul	96
12) Phenanthrene	17.29	178	36282	1.707	ng/ul#	90
13) Anthracene	17.38	178	6670	0.318	ng/ul	92
15) Fluoranthene	19.31	202	21791	0.887	ng/ul	83
17) Pyrene	19.67	202	17927	0.727	ng/ul#	83
18) Benzo(a)anthracene	21.41	228	5461	0.239	ng/ul#	79
19) Chrysene	21.47	228	6412	0.272	ng/ul#	77
21) Benzo(b)fluoranthene	23.19	252	5804m	0.254	ng/ul	
22) Benzo(k)fluoranthene	23.22	252	2649m	0.108	ng/ul	
23) Benzo(a)pyrene	23.86	252	3131m	0.137	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.70	276	2171m	0.092	ng/ul	
26) Benzo(g,h,i)perylene	27.54	276	2035m	0.107	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE110617\
Data File : BE094752.D
Acq On : 6 Nov 2017 14:31
Operator : SJ/JU
Sample : I6150-06
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
B-3(2-3)-102517

Manual Integrations
APPROVED

Sohil
11/7/2017 5:41:32 PM

Quant Time: Nov 07 00:41:06 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE102517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 07 00:12:32 2017
Response via : Initial Calibration

