

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE110617\
 Data File : BE094761.D
 Acq On : 6 Nov 2017 19:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.478

Quant Time: Nov 07 00:15:08 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE102517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 07 00:12:32 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	107	0.00
3	Naphthalene	0.400	0.354	11.5	94	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.366	8.5	97	0.00
5	2-Methylnaphthalene	0.400	0.350	12.5	92	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	96	0.00
7	Acenaphthylene	0.400	0.365	8.8	88	0.00
8 C	Acenaphthene	0.400	0.361	9.8	87	0.00
9	Fluorene	0.400	0.355	11.3	85	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	92	0.00
11	Pentachlorophenol	0.400	0.352	12.0	101	0.00
12	Phenanthrene	0.400	0.355	11.3	82	0.00
13	Anthracene	0.400	0.380	5.0	88	0.00
14 SURR	Fluoranthene-d10	0.400	0.388	3.0	90	0.00
15 C	Fluoranthene	0.400	0.382	4.5	89	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	89	0.00
17	Pyrene	0.400	0.395	1.3	89	0.00
18	Benzo(a)anthracene	0.400	0.391	2.3	89	0.00
19	Chrysene	0.400	0.370	7.5	83	0.00
20 I	Perylene-d12	0.400	0.400	0.0	88	0.00
21	Benzo(b)fluoranthene	0.400	0.386	3.5	86	0.00
22	Benzo(k)fluoranthene	0.400	0.371	7.3	84	0.00
23 C	Benzo(a)pyrene	0.400	0.381	4.8	86	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.334	16.5	75	0.00
25	Dibenzo(a,h)anthracene	0.400	0.356	11.0	80	0.00
26	Benzo(g,h,i)perylene	0.400	0.313	21.8#	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0