

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE110617\
 Data File : BE094767.D
 Acq On : 6 Nov 2017 23:22
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.479

Quant Time: Nov 07 01:53:46 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE102517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 07 01:23:46 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	102	0.00
3	Naphthalene	0.400	0.355	11.3	91	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.368	8.0	93	0.00
5	2-Methylnaphthalene	0.400	0.351	12.3	89	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
7	Acenaphthylene	0.400	0.369	7.8	87	0.00
8 C	Acenaphthene	0.400	0.361	9.8	85	0.00
9	Fluorene	0.400	0.365	8.8	86	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	91	0.00
11	Pentachlorophenol	0.400	0.442	-10.5	126	-0.05
12	Phenanthrene	0.400	0.362	9.5	83	0.00
13	Anthracene	0.400	0.401	-0.3	93	0.00
14 SURR	Fluoranthene-d10	0.400	0.382	4.5	88	0.00
15 C	Fluoranthene	0.400	0.375	6.3	87	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	87	0.00
17	Pyrene	0.400	0.402	-0.5	89	0.00
18	Benzo(a)anthracene	0.400	0.398	0.5	89	0.00
19	Chrysene	0.400	0.358	10.5	79	0.00
20 I	Perylene-d12	0.400	0.400	0.0	84	0.00
21	Benzo(b)fluoranthene	0.400	0.391	2.3	83	0.00
22	Benzo(k)fluoranthene	0.400	0.392	2.0	85	0.00
23 C	Benzo(a)pyrene	0.400	0.394	1.5	85	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.372	7.0	80	0.00
25	Dibenzo(a,h)anthracene	0.400	0.373	6.8	80	0.00
26	Benzo(g,h,i)perylene	0.400	0.382	4.5	83	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0