

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110816\  
 Data File : BE092514.D  
 Acq On : 8 Nov 2016 18:24  
 Operator : UM/SJ  
 Sample : H5504-04  
 Misc :  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 KY038PS033-161031

Quant Time: Nov 09 00:56:09 2016  
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE110816.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Nov 08 16:27:22 2016  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc | Units | Dev(Min)       |
|--------------------------------|-------|------|----------|------|-------|----------------|
| 1) 1,4-Dichlorobenzene-d4      | 8.14  | 152  | 7951     | 5.00 | ng    | -0.02          |
| 7) Naphthalene-d8              | 10.93 | 136  | 35799    | 5.00 | ng    | 0.00           |
| 12) Acenaphthene-d10           | 14.80 | 164  | 23508    | 5.00 | ng    | 0.00           |
| 18) Phenanthrene-d10           | 17.56 | 188  | 52837    | 5.00 | ng    | 0.00           |
| 24) Chrysene-d12               | 21.72 | 240  | 82098    | 5.00 | ng    | 0.00           |
| 31) Perylene-d12               | 24.08 | 264  | 75801    | 5.00 | ng    | -0.01          |
| System Monitoring Compounds    |       |      |          |      |       |                |
| 4) 2-Fluorophenol              | 5.66  | 112  | 6875     | 3.92 | ng    | 0.00           |
| 5) Phenol-d6                   | 7.34  | 99   | 4794     | 2.05 | ng    | 0.00           |
| 8) Nitrobenzene-d5             | 9.32  | 82   | 11127    | 4.74 | ng    | -0.02          |
| 13) 2,4,6-Tribromophenol       | 16.30 | 330  | 5864     | 7.01 | ng    | 0.00           |
| 14) 2-Fluorobiphenyl           | 13.42 | 172  | 27596    | 4.75 | ng    | -0.01          |
| 26) Terphenyl-d14              | 20.16 | 244  | 47694    | 4.78 | ng    | -0.02          |
| Target Compounds               |       |      |          |      |       |                |
| 29) Bis(2-ethylhexyl)phthalate | 21.61 | 149  | 1119     | 0.16 | ng    | Qvalue<br># 77 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110816\  
Data File : BE092514.D  
Acq On : 8 Nov 2016 18:24  
Operator : UM/SJ  
Sample : H5504-04  
Misc :  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
KY038PS033-161031

Quant Time: Nov 09 00:56:09 2016  
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE110816.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Nov 08 16:27:22 2016  
Response via : Initial Calibration

