

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110816\
 Data File : BE092509.D
 Acq On : 8 Nov 2016 15:16
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE110816

Quant Time: Nov 08 16:31:51 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE110816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 16:27:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.775	0.826	-6.6	127	0.00
3	n-Nitrosodimethylamine	0.905	0.751	17.0	100	0.00
4 S	2-Fluorophenol	1.103	1.052	4.6	109	0.00
5 S	Phenol-d6	1.471	1.315	10.6	107	0.00
6	bis(2-Chloroethyl)ether	1.610	1.465	9.0	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.328	0.304	7.3	111	0.00
9	Naphthalene	1.082	1.070	1.1	102	0.00
10	Hexachlorobutadiene	0.164	0.169	-3.0	104	-0.02
11	2-Methylnaphthalene	0.703	0.687	2.3	106	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
13 S	2,4,6-Tribromophenol	0.142	0.111	21.8	106	0.00
14 S	2-Fluorobiphenyl	1.237	1.263	-2.1	111	0.00
15	Acenaphthylene	7.172	6.975	2.7	109	-0.02
16	Acenaphthene	1.159	1.143	1.4	107	0.00
17	Fluorene	1.455	1.422	2.3	108	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
19	Hexachlorobenzene	0.189	0.193	-2.1	112	0.00
20	Pentachlorophenol	0.055	0.044	20.0	97	0.00
21	Phenanthrene	1.081	1.156	-6.9	113	0.00
22	Anthracene	1.048	1.048	0.0	107	0.00
23	Fluoranthene	5.427	5.455	-0.5	106	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
25	Pyrene	4.097	3.757	8.3	107	0.00
26 S	Terphenyl-d14	0.607	0.566	6.8	109	-0.02
27	Benzo(a)anthracene	5.068	4.809	5.1	111	0.00
28	Chrysene	5.430	5.192	4.4	101	0.00
29	Bis(2-ethylhexyl)phthalate	0.535	0.445	16.8	102	0.00
30	Indeno(1,2,3-cd)pyrene	5.531	5.280	4.5	102	0.00
31 I	Perylene-d12	1.000	1.000	0.0	103	0.00
32	Benzo(b)fluoranthene	1.885	1.959	-3.9	103	0.00
33	Benzo(k)fluoranthene	1.972	1.824	7.5	97	0.00
34 C	Benzo(a)pyrene	1.292	1.233	4.6	102	0.00
35	Dibenzo(a,h)anthracene	1.216	1.174	3.5	102	-0.02
36	Benzo(g,h,i)perylene	5.009	4.857	3.0	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0