

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110816\
 Data File : BE092517.D
 Acq On : 8 Nov 2016 20:15
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 09 09:58:25 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE110816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 16:27:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	-0.02
2	1,4-Dioxane	0.775	0.774	0.1	127	0.00
3	n-Nitrosodimethylamine	0.905	0.655	27.6#	93	0.01
4 S	2-Fluorophenol	1.103	1.018	7.7	112	0.00
5 S	Phenol-d6	1.471	1.213	17.5	105	0.00
6	bis(2-Chloroethyl)ether	1.610	1.349	16.2	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.328	0.290	11.6	115	0.00
9	Naphthalene	1.082	1.046	3.3	108	0.00
10	Hexachlorobutadiene	0.164	0.162	1.2	108	-0.02
11	2-Methylnaphthalene	0.703	0.663	5.7	111	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
13 S	2,4,6-Tribromophenol	0.142	0.116	18.3	117	0.00
14 S	2-Fluorobiphenyl	1.237	1.262	-2.0	118	0.00
15	Acenaphthylene	7.172	7.003	2.4	117	-0.02
16	Acenaphthene	1.159	1.137	1.9	114	0.00
17	Fluorene	1.455	1.434	1.4	116	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
19	Hexachlorobenzene	0.189	0.186	1.6	115	0.00
20	Pentachlorophenol	0.055	0.043	21.8	100	0.00
21	Phenanthrene	1.081	1.117	-3.3	116	0.00
22	Anthracene	1.048	1.043	0.5	112	0.00
23	Fluoranthene	5.427	5.249	3.3	108	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
25	Pyrene	4.097	3.888	5.1	111	0.00
26 S	Terphenyl-d14	0.607	0.591	2.6	114	0.00
27	Benzo(a)anthracene	5.068	4.858	4.1	112	0.00
28	Chrysene	5.430	5.104	6.0	99	-0.02
29	Bis(2-ethylhexyl)phthalate	0.535	0.430	19.6	99	-0.02
30	Indeno(1,2,3-cd)pyrene	5.531	5.186	6.2	100	0.00
31 I	Perylene-d12	1.000	1.000	0.0	101	-0.01
32	Benzo(b)fluoranthene	1.885	1.867	1.0	97	0.00
33	Benzo(k)fluoranthene	1.972	1.695	14.0	89	0.00
34 C	Benzo(a)pyrene	1.292	1.198	7.3	97	0.00
35	Dibenzo(a,h)anthracene	1.216	1.151	5.3	98	-0.02
36	Benzo(g,h,i)perylene	5.009	4.846	3.3	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0