

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110816\
 Data File : BE092517.D
 Acq On : 8 Nov 2016 20:15
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 09 09:58:25 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE110816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 16:27:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	112	-0.02
2	1,4-Dioxane	0.400	0.390	2.5	127	0.00
3	n-Nitrosodimethylamine	0.400	0.369	7.8	93	0.01
4 S	2-Fluorophenol	0.400	0.369	7.8	112	0.00
5 S	Phenol-d6	0.400	0.330	17.5	105	0.00
6	bis(2-Chloroethyl)ether	0.400	0.335	16.3	102	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.400	0.354	11.5	115	0.00
9	Naphthalene	0.400	0.387	3.3	108	0.00
10	Hexachlorobutadiene	0.400	0.394	1.5	108	-0.02
11	2-Methylnaphthalene	0.400	0.377	5.8	111	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	112	0.00
13 S	2,4,6-Tribromophenol	0.400	0.480	-20.0	117	0.00
14 S	2-Fluorobiphenyl	0.400	0.408	-2.0	118	0.00
15	Acenaphthylene	0.400	0.391	2.3	117	-0.02
16	Acenaphthene	0.400	0.392	2.0	114	0.00
17	Fluorene	0.400	0.394	1.5	116	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	109	0.00
19	Hexachlorobenzene	0.400	0.394	1.5	115	0.00
20	Pentachlorophenol	0.400	0.316	21.0	100	0.00
21	Phenanthrene	0.400	0.413	-3.2	116	0.00
22	Anthracene	0.400	0.398	0.5	112	0.00
23	Fluoranthene	0.400	0.387	3.3	108	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	108	0.00
25	Pyrene	0.400	0.380	5.0	111	0.00
26 S	Terphenyl-d14	0.400	0.389	2.8	114	0.00
27	Benzo(a)anthracene	0.400	0.383	4.3	112	0.00
28	Chrysene	0.400	0.348	13.0	99	-0.02
29	Bis(2-ethylhexyl)phthalate	0.400	0.393	1.8	99	-0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.375	6.3	100	0.00
31 I	Perylene-d12	5.000	5.000	0.0	101	-0.01
32	Benzo(b)fluoranthene	0.400	0.393	1.8	97	0.00
33	Benzo(k)fluoranthene	0.400	0.319	20.3	89	0.00
34 C	Benzo(a)pyrene	0.400	0.371	7.3	97	0.00
35	Dibenzo(a,h)anthracene	0.400	0.379	5.3	98	-0.02
36	Benzo(g,h,i)perylene	0.400	0.387	3.3	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0