

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110816\
 Data File : BE092507.D
 Acq On : 8 Nov 2016 13:59
 Operator : UM/SJ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED

umangi
 11/9/2016 10:58:50 AM

Quant Time: Nov 08 15:29:11 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE110816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 12:48:28 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	7951	5.00	ng	-0.02
7) Naphthalene-d8	10.93	136	40807	5.00	ng	0.00
12) Acenaphthene-d10	14.80	164	29444	5.00	ng	0.00
18) Phenanthrene-d10	17.56	188	65408	5.00	ng	0.00
24) Chrysene-d12	21.72	240	91110	5.00	ng	0.00
31) Perylene-d12	24.08	264	82242	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.65	112	9471	7.86	ng	-0.01
5) Phenol-d6	7.31	99	13216	7.84	ng	-0.02
8) Nitrobenzene-d5	9.32	82	14465	6.31	ng	-0.02
13) 2,4,6-Tribromophenol	16.30	330	5007	8.67	ng	-0.01
14) 2-Fluorobiphenyl	13.42	172	36472	5.08	ng	-0.01
26) Terphenyl-d14	20.16	244	57752	5.12	ng	-0.02

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.39	88	4818	4.63	ng	94
3) n-Nitrosodimethylamine	3.74	42	7023	7.36	ng	99
6) bis(2-Chloroethyl)ether	7.57	93	13401	6.74	ng	89
9) Naphthalene	10.97	128	41701	4.73	ng	# 95
10) Hexachlorobutadiene	11.26	225	6287	4.65	ng	99
11) 2-Methylnaphthalene	12.60	142	28515	5.00	ng	98
15) Acenaphthylene	14.51	152	217104	5.16	ng	100
16) Acenaphthene	14.87	154	32586	4.85	ng	94
17) Fluorene	15.84	166	43043	5.35	ng	99
19) Hexachlorobenzene	16.87	284	12455	4.26	ng	96
21) Phenanthrene	17.58	178	69312	4.10	ng	# 78
22) Anthracene	17.67	178	69515	4.57	ng	99
23) Fluoranthene	19.59	202	367362	5.48	ng	99
25) Pyrene	19.95	202	382059	4.81	ng	99
27) Benzo(a)anthracene	21.70	228	450052m	5.34	ng	
28) Chrysene	21.75	228	381334m	4.30	ng	
29) Bis(2-ethylhexyl)phthalate	21.63	149	40863	5.82	ng	# 72
30) Indeno(1,2,3-cd)pyrene	26.56	276	478348	6.13	ng	99
32) Benzo(b)fluoranthene	23.35	252	106922	5.22	ng	# 95
33) Benzo(k)fluoranthene	23.39	252	111590	5.50	ng	93
34) Benzo(a)pyrene	23.97	252	100987	5.45	ng	# 95
35) Dibenzo(a,h)anthracene	26.58	278	98363	5.68	ng	94
36) Benzo(g,h,i)perylene	27.31	276	402293	5.39	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE110816\
Data File : BE092507.D
Acq On : 8 Nov 2016 13:59
Operator : UM/SJ
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

umangi
11/9/2016 10:58:50 AM

Quant Time: Nov 08 15:29:11 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE110816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 08 12:48:28 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.14 min Scan# 575

Delta R.T. -0.02 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion:	152	Resp:	7951
Ion Ratio	Lower	Upper	
152	100		
150	188.0	106.0	159.0#
115	96.2	31.2	46.8#

Manual Integrations
APPROVED

umangi
11/9/2016 10:58:50 AM



#2

1,4-Dioxane

Concen: 4.63 ng




RT: 3.39 min Scan# 69


Delta R.T. 0.00 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Tgt Ion:	88	Resp:	4818
Ion Ratio	Lower	Upper	
88	100		
58	86.0	63.6	95.4
43	36.6	28.0	42.0









#3
n-Nitrosodimethylamine
Concen: 7.36 ng
RT: 3.74 min Scan# 99
Delta R.T. -0.02 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
42	100		
74	108.3	86.0	129.0
44	6.4	5.1	7.7




Manual Integrations
APPROVED


umangi
11/9/2016 10:58:50 AM



#4
2-Fluorophenol
Concen: 7.86 ng
RT: 5.65 min Scan# 360
Delta R.T. -0.01 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
112	100		
64	66.6	53.5	80.3
63	35.7	27.9	41.9





#5

Phenol-d6

Concen: 7.84 ng

RT: 7.31 min Scan# 539

Delta R.T. -0.02 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion:	99	Resp:	13216
Ion Ratio	Lower	Upper	
99	100		
42	24.6	16.0	24.0#
71	36.7	30.6	45.8

Manual Integrations
APPROVED

umangi
11/9/2016 10:58:50 AM



#6

bis(2-Chloroethyl)ether

Concen: 6.74 ng




RT: 7.57 min Scan# 550


Delta R.T. -0.02 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Tgt Ion:	93	Resp:	13401
Ion Ratio	Lower	Upper	
93	100		
63	76.2	71.6	107.4
95	31.6	24.0	36.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.93 min Scan# 707

Delta R.T. 0.00 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Instrument :

BNA_E


ClientSampleId :

SSTDIC005

Tot Ion	Ratio	Resp	Lower	Upper
136	100			
137	11.2	8.2	12.4	
54	6.2	9.6	14.4	
68	4.9	6.7	10.1	

Manual Integrations
APPROVED

umangi
11/9/2016 10:58:50 AM






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 6.31 ng


RT: 9.32 min Scan# 626

Delta R.T. -0.02 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59



Tgt Ion	Ratio	Resp	Lower	Upper
82	100			
128	31.6	39.0	58.4	
54	47.9	44.8	67.2	




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 4.73 ng

RT: 10.97 min Scan# 709

Delta R.T. -0.02 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Instrument :

BNA_E





ClientSampleId :

SSTDICC005

Tot Ion:	128	Resp:	41701
Ion Ratio	Lower	Upper	
128	100		
129	10.3	11.2	16.8#
127	14.2	11.6	17.4

Manual Integrations
APPROVED

umangi
11/9/2016 10:58:50 AM



#10

Hexachlorobutadiene

Concen: 4.65 ng




RT: 11.26 min Scan# 724


Delta R.T. -0.02 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Tgt Ion:	225	Resp:	6287
Ion Ratio	Lower	Upper	
225	100		
223	62.8	50.6	76.0
227	63.8	51.4	77.0









#11
2-Methylnaphthalene
Concen: 5.00 ng
RT: 12.60 min Scan# 812
Delta R.T. -0.02 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
142	100		
141	92.7	73.7	110.5
115	39.2	34.0	51.0




Manual Integrations
APPROVED


umangi
11/9/2016 10:58:50 AM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.80 min Scan# 978
Delta R.T. -0.00 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
164	100		
162	107.4	71.4	107.0#
160	48.5	27.4	41.2#









#13
 2,4,6-Tribromophenol
 Concen: 8.67 ng
 RT: 16.30 min Scan# 1086
 Delta R.T. -0.01 min
 Lab File: BE092507.D
 Acq: 8 Nov 2016 13:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Tot Ion:330 Resp: 5007
 Ion Ratio Lower Upper
 330 100
 332 95.2 75.4 113.0
 141 40.8 31.0 46.6




Manual Integrations
 APPROVED


umangi
 11/9/2016 10:58:50 AM



#14
 2-Fluorobiphenyl
 Concen: 5.08 ng
 RT: 13.42 min Scan# 883
 Delta R.T. -0.01 min
 Lab File: BE092507.D
 Acq: 8 Nov 2016 13:59

Tgt Ion:172 Resp: 36472
 Ion Ratio Lower Upper
 172 100
 171 31.0 30.1 45.1
 170 19.0 21.8 32.6#





#15

Acenaphthylene

Concen: 5.16 ng

RT: 14.51 min Scan# 961

Delta R.T. -0.02 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.9	5.9
153	13.1	10.4	15.6

Manual Integrations
APPROVED

umangi
11/9/2016 10:58:50 AM



#16

Acenaphthene

Concen: 4.85 ng




RT: 14.87 min Scan# 982


Delta R.T. -0.00 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
154	100		
153	452.4	350.8	526.2
152	223.7	171.1	256.7









#17
Fluorene
Concen: 5.35 ng
RT: 15.84 min Scan# 1047
Delta R.T. -0.01 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
166	100		
165	98.8	78.2	117.4
167	13.5	11.0	16.4




Manual Integrations
APPROVED


umangi
11/9/2016 10:58:50 AM



#18
Phenanthrene-d10
Concen: 5.00 ng
RT: 17.56 min Scan# 1152
Delta R.T. -0.00 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
188	100		
94	3.8	3.2	4.8
80	3.2	2.6	3.8









#19
Hexachlorobenzene
Concen: 4.26 ng
RT: 16.87 min Scan# 1122
Delta R.T. -0.00 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
284	100		
142	38.6	29.1	43.7
249	36.9	27.9	41.9




Manual Integrations
APPROVED


umangi
11/9/2016 10:58:50 AM



#21
Phenanthrene
Concen: 4.10 ng
RT: 17.58 min Scan# 1153
Delta R.T. -0.02 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
178	100		
176	20.1	15.8	23.8
179	0.0	16.0	24.0#





#22

Anthracene

Concen: 4.57 ng

RT: 17.67 min Scan# 1157

Delta R.T. -0.02 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion	Ratio	Lower	Upper
178	100		
176	19.7	15.0	22.4
179	15.0	12.2	18.2

Manual Integrations
APPROVED

umangi
11/9/2016 10:58:50 AM



#23

Fluoranthene

Concen: 5.48 ng




RT: 19.59 min Scan# 1264


Delta R.T. 0.00 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
202	100		
101	2.7	2.2	3.4
203	17.5	13.7	20.5









#24
Chrysene-d12
Concen: 5.00 ng
RT: 21.72 min Scan# 1373
Delta R.T. -0.00 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
240	100		
120	2.9	2.6	4.0
236	29.8	24.6	37.0




Manual Integrations
APPROVED


umangi
11/9/2016 10:58:50 AM



#25
Pyrene
Concen: 4.81 ng
RT: 19.95 min Scan# 1285
Delta R.T. 0.00 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
202	100		
200	5.3	4.2	6.4
203	18.2	13.9	20.9









#26
 Terphenyl-d14
 Concen: 5.12 ng
 RT: 20.16 min Scan# 1297
 Delta R.T. -0.02 min
 Lab File: BE092507.D
 Acq: 8 Nov 2016 13:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Tot Ion	Ratio	Lower	Upper
244	100		
212	8.6	6.7	10.1
122	5.5	3.6	5.4




Manual Integrations
 APPROVED


umangi
 11/9/2016 10:58:50 AM



#27
 Benzo(a)anthracene
 Concen: 5.34 ng m
 RT: 21.70 min Scan# 1372
 Delta R.T. -0.00 min
 Lab File: BE092507.D
 Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.3	23.6	35.4
229	16.2	13.3	19.9









#28
Chrysene
Concen: 4.30 ng/ml
RT: 21.75 min Scan# 1374
Delta R.T. -0.02 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
228	100		
226	45.6	36.3	54.5
229	11.5	10.6	16.0




Manual Integrations
APPROVED


umangi
11/9/2016 10:58:50 AM



#29
Bis(2-ethylhexyl)phthalate
Concen: 5.82 ng/ml
RT: 21.63 min Scan# 1369
Delta R.T. -0.00 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
149	100		
167	29.0	16.0	24.0#
279	3.6	16.0	24.0#









#30
Indeno(1,2,3-cd)pyrene
Concen: 6.13 ng
RT: 26.56 min Scan# 1713
Delta R.T. -0.00 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
276	100		
138	26.6	20.3	30.5
277	24.8	19.7	29.5




Manual Integrations
APPROVED


umangi
11/9/2016 10:58:50 AM



#31
Perylene-d12
Concen: 5.00 ng
RT: 24.08 min Scan# 1542
Delta R.T. -0.01 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
264	100		
260	26.6	20.1	30.1
265	21.1	17.9	26.9





#32

Benzo(b)fluoranthene

Concen: 5.22 ng

RT: 23.35 min Scan# 1479

Delta R.T. -0.01 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Instrument :

BNA_E





ClientSampleId :

SSTDIC005

Tot Ion:	252	Resp:	106922
Ion Ratio	Lower	Upper	
252	100		
253	21.4	18.7	28.1
125	10.4	10.5	15.7

Manual Integrations
APPROVED

umangi
11/9/2016 10:58:50 AM



#33

Benzo(k)fluoranthene

Concen: 5.50 ng




RT: 23.39 min Scan# 1483


Delta R.T. -0.01 min

Lab File: BE092507.D

Acq: 8 Nov 2016 13:59

Tgt Ion:	252	Resp:	111590
Ion Ratio	Lower	Upper	
252	100		
253	21.0	20.3	30.5
125	10.4	9.6	14.4









#34
Benzo(a)pyrene
Concen: 5.45 ng
RT: 23.97 min Scan# 1533
Delta R.T. 0.00 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.3	19.0	28.6
125	11.0	11.8	17.8




Manual Integrations
APPROVED




umangi
11/9/2016 10:58:50 AM



#35
Dibenzo(a,h)anthracene
Concen: 5.68 ng
RT: 26.58 min Scan# 1714
Delta R.T. -0.02 min
Lab File: BE092507.D
Acq: 8 Nov 2016 13:59

Tgt Ion	Ratio	Lower	Upper
278	100		
139	14.3	13.8	20.8
279	24.2	21.4	32.2





#36
 Benzo(a,h,i)perylene
 Concen: 5.39 ng
 RT: 27.31 min Scan# 1757
 Delta R.T. -0.02 min
 Lab File: BE092507.D
 Acq: 8 Nov 2016 13:59

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Tot Ion:	276	Resp:	402293
Ion Ratio	Lower	Upper	
276	100		
277	22.2	19.8	29.8
138	24.0	19.8	29.6

Manual Integrations
 APPROVED

umangi
 11/9/2016 10:58:50 AM

