

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110916\
 Data File : BE092519.D
 Acq On : 9 Nov 2016 10:58
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 09 14:29:43 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE110816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 16:27:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	-0.02
2	1,4-Dioxane	0.775	0.782	-0.9	131	0.00
3	n-Nitrosodimethylamine	0.905	0.665	26.5#	96	0.00
4 S	2-Fluorophenol	1.103	0.972	11.9	109	0.00
5 S	Phenol-d6	1.471	1.200	18.4	106	0.00
6	bis(2-Chloroethyl)ether	1.610	1.331	17.3	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.328	0.290	11.6	116	0.00
9	Naphthalene	1.082	1.033	4.5	108	0.00
10	Hexachlorobutadiene	0.164	0.161	1.8	109	-0.02
11	2-Methylnaphthalene	0.703	0.659	6.3	111	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
13 S	2,4,6-Tribromophenol	0.142	0.107	24.6	110	0.00
14 S	2-Fluorobiphenyl	1.237	1.278	-3.3	121	0.00
15	Acenaphthylene	7.172	6.997	2.4	118	-0.02
16	Acenaphthene	1.159	1.146	1.1	116	0.00
17	Fluorene	1.455	1.428	1.9	117	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
19	Hexachlorobenzene	0.189	0.169	10.6	112	0.00
20	Pentachlorophenol	0.055	0.044	20.0	110	0.00
21	Phenanthrene	1.081	1.050	2.9	117	0.00
22	Anthracene	1.048	1.001	4.5	116	0.00
23	Fluoranthene	5.427	5.034	7.2	111	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
25	Pyrene	4.097	3.723	9.1	113	0.00
26 S	Terphenyl-d14	0.607	0.555	8.6	114	0.00
27	Benzo(a)anthracene	5.068	4.543	10.4	111	0.00
28	Chrysene	5.430	4.813	11.4	99	-0.02
29	Bis(2-ethylhexyl)phthalate	0.535	0.350	34.6#	85	-0.02
30	Indeno(1,2,3-cd)pyrene	5.531	5.151	6.9	105	0.00
31 I	Perylene-d12	1.000	1.000	0.0	107	0.00
32	Benzo(b)fluoranthene	1.885	1.736	7.9	95	0.00
33	Benzo(k)fluoranthene	1.972	1.765	10.5	97	0.00
34 C	Benzo(a)pyrene	1.292	1.152	10.8	98	0.00
35	Dibenzo(a,h)anthracene	1.216	1.166	4.1	105	-0.02
36	Benzo(g,h,i)perylene	5.009	4.869	2.8	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0