

Data Path : Z:\HPCHEM1\BNA E\DATA\BE110916\
 Data File : BE092525.D
 Acq On : 9 Nov 2016 15:02
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 09 15:51:23 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE110816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 16:27:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	115	-0.02
2	1,4-Dioxane	0.400	0.328	18.0	116	0.00
3	n-Nitrosodimethylamine	0.400	0.385	3.8	101	0.01
4 S	2-Fluorophenol	0.400	0.363	9.3	114	0.00
5 S	Phenol-d6	0.400	0.342	14.5	112	0.00
6	bis(2-Chloroethyl)ether	0.400	0.336	16.0	105	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	118	0.00
8 S	Nitrobenzene-d5	0.400	0.367	8.3	124	0.00
9	Naphthalene	0.400	0.389	2.8	114	0.00
10	Hexachlorobutadiene	0.400	0.393	1.8	113	-0.02
11	2-Methylnaphthalene	0.400	0.381	4.8	117	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	120	0.00
13 S	2,4,6-Tribromophenol	0.400	0.480	-20.0	126	-0.01
14 S	2-Fluorobiphenyl	0.400	0.407	-1.7	126	0.00
15	Acenaphthylene	0.400	0.390	2.5	125	-0.02
16	Acenaphthene	0.400	0.395	1.3	122	0.00
17	Fluorene	0.400	0.394	1.5	124	-0.01
18 I	Phenanthrene-d10	5.000	5.000	0.0	115	0.00
19	Hexachlorobenzene	0.400	0.435	-8.7	133	0.00
20	Pentachlorophenol	0.400	0.339	15.3	113	0.00
21	Phenanthrene	0.400	0.419	-4.7	123	-0.02
22	Anthracene	0.400	0.403	-0.8	119	0.00
23	Fluoranthene	0.400	0.401	-0.3	118	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	122	0.00
25	Pyrene	0.400	0.373	6.8	123	0.00
26 S	Terphenyl-d14	0.400	0.358	10.5	119	-0.02
27	Benzo(a)anthracene	0.400	0.363	9.3	120	0.00
28	Chrysene	0.400	0.313	21.8	105	-0.02
29	Bis(2-ethylhexyl)phthalate	0.400	0.320	20.0	91	-0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.351	12.3	106	0.00
31 I	Perylene-d12	5.000	5.000	0.0	109	-0.01
32	Benzo(b)fluoranthene	0.400	0.323	19.3	92	-0.01
33	Benzo(k)fluoranthene	0.400	0.227	43.3#	78	0.00
34 C	Benzo(a)pyrene	0.400	0.359	10.3	101	0.00
35	Dibenzo(a,h)anthracene	0.400	0.374	6.5	104	-0.02
36	Benzo(g,h,i)perylene	0.400	0.383	4.3	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0