

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE110917\
 Data File : BE094785.D
 Acq On : 8 Nov 2017 16:33
 Operator : SJ/JU
 Sample : PB103996BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB103996BL

Manual Integrations
 APPROVED

Sohil
 11/9/2017 3:45:13 PM

Quant Time: Nov 09 00:36:54 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.93	152	1028	0.40	ng	0.02
7) Naphthalene-d8	10.76	136	6403	0.40	ng	0.05
13) Acenaphthene-d10	14.54	164	3987	0.40	ng	0.03
19) Phenanthrene-d10	17.27	188	8795	0.40	ng	0.03
27) Chrysene-d12	21.45	240	9184	0.40	ng	0.02
34) Perylene-d12	23.99	264	8328	0.40	ng	0.04
System Monitoring Compounds						
4) 2-Fluorophenol	5.56	112	748	0.21	ng	0.04
5) Phenol-d6	7.33	99	1089	0.20	ng	0.19
8) Nitrobenzene-d5	9.21	82	829	0.16	ng	0.13
11) 2-Methylnaphthalene-d10	12.36	152	3122	0.30	ng	0.07
14) 2,4,6-Tribromophenol	16.10	330	256m	0.23	ng	0.07
15) 2-Fluorobiphenyl	13.21	172	3701	0.26	ng	0.06
25) Fluoranthene-d10	19.30	212	31172	0.26	ng	0.02
29) Terphenyl-d14	19.87	244	5153	0.30	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE110917\
Data File : BE094785.D
Acq On : 8 Nov 2017 16:33
Operator : SJ/JU
Sample : PB103996BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB103996BL

Manual Integrations
APPROVED

Sohil
11/9/2017 3:45:13 PM

Quant Time: Nov 09 00:36:54 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

