

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE110917\
 Data File : BE094784.D
 Acq On : 8 Nov 2017 15:43
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 09 00:29:30 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.01
2	1,4-Dioxane	0.677	0.618	8.7	78	0.00
3	n-Nitrosodimethylamine	0.648	0.469	27.6#	71	0.02
4 S	2-Fluorophenol	1.370	1.161	15.3	84	0.05
5 S	Phenol-d6	2.069	1.989	3.9	97	0.06
6	bis(2-Chloroethyl)ether	1.600	1.439	10.1	90	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.319	0.259	18.8	91	0.05
9	Naphthalene	4.527	4.008	11.5	92	0.02
10	Hexachlorobutadiene	0.164	0.164	0.0	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.642	0.570	11.2	93	0.02
12	2-Methylnaphthalene	0.768	0.661	13.9	91	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.02
14 S	2,4,6-Tribromophenol	0.110	0.095	13.6	85	0.03
15 S	2-Fluorobiphenyl	1.410	1.329	5.7	88	0.02
16	Acenaphthylene	2.112	1.934	8.4	87	0.00
17	Acenaphthene	1.356	1.249	7.9	86	0.00
18	Fluorene	1.738	1.552	10.7	85	0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.03
20	4-Bromophenyl-phenylether	0.218	0.156	28.4#	80	0.03
21	Hexachlorobenzene	0.191	0.232	-21.5	116	0.00
22	Pentachlorophenol	0.024	0.022	8.3	98	0.07
23	Phenanthrene	1.141	1.206	-5.7	115	0.00
24	Anthracene	1.176	1.119	4.8	102	0.00
25 SURR	Fluoranthene-d10	4.676	4.360	6.8	89	0.01
26	Fluoranthene	1.420	1.296	8.7	87	0.01
27 I	Chrysene-d12	1.000	1.000	0.0	90	0.01
28	Pvrene	1.399	1.340	4.2	88	0.01
29 S	Terphenyl-d14	0.754	0.704	6.6	86	0.00
30	Benzo(a)anthracene	1.302	1.219	6.4	85	0.01
31	Chrysene	1.322	1.242	6.1	88	0.01
32	Bis(2-ethylhexyl)phthalate	3.193	3.029	5.1	94	-0.01
33	Indeno(1,2,3-cd)pyrene	1.222	0.952	22.1	72	0.07
34 I	Perylene-d12	1.000	1.000	0.0	85	0.03
35	Benzo(b)fluoranthene	1.294	1.156	10.7	78	0.02
36	Benzo(k)fluoranthene	1.379	1.379	0.0	86	0.03
37 C	Benzo(a)pyrene	1.255	1.215	3.2	83	0.04
38	Dibenzo(a,h)anthracene	1.127	0.907	19.5	70	0.05
39	Benzo(a,h,i)perylene	1.055	0.712	32.5#	58	0.09

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE110917\
Data File : BE094784.D
Acq On : 8 Nov 2017 15:43
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 09 00:29:30 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			