

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE110917\
 Data File : BE094784.D
 Acq On : 8 Nov 2017 15:43
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 09 00:29:30 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.01
2	1,4-Dioxane	0.400	0.365	8.8	78	0.00
3	n-Nitrosodimethylamine	0.400	0.290	27.5#	71	0.02
4 S	2-Fluorophenol	0.400	0.339	15.3	84	0.05
5 S	Phenol-d6	0.400	0.385	3.8	97	0.06
6	bis(2-Chloroethyl)ether	0.400	0.360	10.0	90	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	107	0.00
8 S	Nitrobenzene-d5	0.400	0.325	18.8	91	0.05
9	Naphthalene	0.400	0.354	11.5	92	0.02
10	Hexachlorobutadiene	0.400	0.400	0.0	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.355	11.3	93	0.02
12	2-Methylnaphthalene	0.400	0.344	14.0	91	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	0.02
14 S	2,4,6-Tribromophenol	0.400	0.345	13.8	85	0.03
15 S	2-Fluorobiphenyl	0.400	0.377	5.8	88	0.02
16	Acenaphthylene	0.400	0.366	8.5	87	0.00
17	Acenaphthene	0.400	0.369	7.8	86	0.00
18	Fluorene	0.400	0.357	10.8	85	0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.03
20	4-Bromophenyl-phenylether	0.400	0.286	28.5#	80	0.03
21	Hexachlorobenzene	0.400	0.485	-21.2	116	0.00
22	Pentachlorophenol	0.400	0.356	11.0	98	0.07
23	Phenanthrene	0.400	0.462	-15.5	115	0.00
24	Anthracene	0.400	0.381	4.8	102	0.00
25 SURR	Fluoranthene-d10	0.400	0.343	14.2	89	0.01
26	Fluoranthene	0.400	0.333	16.8	87	0.01
27 I	Chrysene-d12	0.400	0.400	0.0	90	0.01
28	Pvrene	0.400	0.383	4.3	88	0.01
29 S	Terphenyl-d14	0.400	0.374	6.5	86	0.00
30	Benzo(a)anthracene	0.400	0.374	6.5	85	0.01
31	Chrysene	0.400	0.376	6.0	88	0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.379	5.3	94	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.312	22.0	72	0.07
34 I	Perylene-d12	0.400	0.400	0.0	85	0.03
35	Benzo(b)fluoranthene	0.400	0.357	10.8	78	0.02
36	Benzo(k)fluoranthene	0.400	0.400	0.0	86	0.03
37 C	Benzo(a)pyrene	0.400	0.387	3.3	83	0.04
38	Dibenzo(a,h)anthracene	0.400	0.322	19.5	70	0.05
39	Benzo(a,h,i)perylene	0.400	0.270	32.5#	58	0.09

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE110917\
Data File : BE094784.D
Acq On : 8 Nov 2017 15:43
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 09 00:29:30 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			