

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE110917\
 Data File : BE094793.D
 Acq On : 8 Nov 2017 21:17
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 09 00:50:07 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 22:40:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.01
2	1,4-Dioxane	0.677	0.603	10.9	78	-0.01
3	n-Nitrosodimethylamine	0.648	0.471	27.3#	73	0.03
4 S	2-Fluorophenol	1.370	1.027	25.0#	76	0.05
5 S	Phenol-d6	2.069	1.902	8.1	96	0.06
6	bis(2-Chloroethyl)ether	1.600	1.405	12.2	90	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.319	0.260	18.5	92	0.05
9	Naphthalene	4.527	3.993	11.8	93	0.02
10	Hexachlorobutadiene	0.164	0.163	0.6	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.642	0.571	11.1	94	0.02
12	2-Methylnaphthalene	0.768	0.661	13.9	92	0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.02
14 S	2,4,6-Tribromophenol	0.110	0.100	9.1	91	0.03
15 S	2-Fluorobiphenyl	1.410	1.314	6.8	89	0.02
16	Acenaphthylene	2.112	1.936	8.3	89	0.00
17	Acenaphthene	1.356	1.239	8.6	87	0.00
18	Fluorene	1.738	1.544	11.2	86	0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.03
20	4-Bromophenyl-phenylether	0.218	0.161	26.1#	81	0.03
21	Hexachlorobenzene	0.191	0.230	-20.4	113	0.00
22	Pentachlorophenol	0.024	0.027	-12.5	118	0.03
23	Phenanthrene	1.141	1.243	-8.9	117	0.00
24	Anthracene	1.176	1.164	1.0	105	0.00
25 SURR	Fluoranthene-d10	4.676	4.413	5.6	89	0.01
26	Fluoranthene	1.420	1.332	6.2	88	0.01
27 I	Chrysene-d12	1.000	1.000	0.0	92	0.01
28	Pvrene	1.399	1.326	5.2	89	0.01
29 S	Terphenyl-d14	0.754	0.692	8.2	86	0.00
30	Benzo(a)anthracene	1.302	1.225	5.9	86	0.01
31	Chrysene	1.322	1.239	6.3	89	0.01
32	Bis(2-ethylhexyl)phthalate	3.193	2.920	8.5	92	-0.01
33	Indeno(1,2,3-cd)pyrene	1.222	0.813	33.5#	62	0.07
34 I	Perylene-d12	1.000	1.000	0.0	85	0.03
35	Benzo(b)fluoranthene	1.294	1.193	7.8	80	0.02
36	Benzo(k)fluoranthene	1.379	1.470	-6.6	91	0.02
37 C	Benzo(a)pyrene	1.255	1.211	3.5	82	0.03
38	Dibenzo(a,h)anthracene	1.127	0.893	20.8	69	0.05
39	Benzo(a,h,i)perylene	1.055	0.690	34.6#	56	0.09

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE110917\
Data File : BE094793.D
Acq On : 8 Nov 2017 21:17
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 09 00:50:07 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE102317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 22:40:47 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			