

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111015\
 Data File : BE091099.D
 Acq On : 10 Nov 2015 17:43
 Operator : UM/NP
 Sample : G4308-05
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 D9KR4

Manual Integrations
 APPROVED

Mmdadoda
 11/13/2015 6:28:24 PM

Quant Time: Nov 13 14:23:40 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 13 13:56:00 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	3529	0.40	ng/ul	0.00
4) Naphthalene-d8	9.57	136	20437	0.40	ng/ul	0.00
8) Acenaphthene-d10	13.40	164	12234	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.14	188	28634m	0.40	ng/ul	0.00
18) Chrysene-d12	20.29	240	35312	0.40	ng/ul	0.00
22) Perylene-d12	22.19	264	31200m	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.95	96	5939	2.34	ng/uL	0.00
6) 2-Methylnaphthalene-d10	11.08	152	8037	0.29	ng/ul	0.00
16) Fluoranthene-d10	18.12	212	23116	0.29	ng/ul	0.00
Target Compounds						
5) Naphthalene	9.61	128	2356	0.05	ng/ul#	Ovalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111015\
Data File : BE091099.D
Acq On : 10 Nov 2015 17:43
Operator : UM/NP
Sample : G4308-05
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
D9KR4

Manual Integrations
APPROVED

Mmdadoda
11/13/2015 6:28:24 PM

Quant Time: Nov 13 14:23:40 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 13 13:56:00 2015
Response via : Initial Calibration

