

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111015\
 Data File : BE091111.D
 Acq On : 11 Nov 2015 00:59
 Operator : UM/NP
 Sample : G4308-17
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 D9KR3

Manual Integrations
 APPROVED

Mmdadoda
 11/13/2015 6:29:45 PM

Quant Time: Nov 13 15:12:12 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 13 13:56:00 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	3453	0.40	ng/ul	0.00
4) Naphthalene-d8	9.57	136	19784	0.40	ng/ul	0.00
8) Acenaphthene-d10	13.40	164	12065	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.14	188	29106m	0.40	ng/ul	0.00
18) Chrysene-d12	20.29	240	34189	0.40	ng/ul	0.00
22) Perylene-d12	22.19	264	30220m	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.95	96	4999	2.01	ng/uL	0.00
6) 2-Methylnaphthalene-d10	11.08	152	7676	0.28	ng/ul	0.00
16) Fluoranthene-d10	18.12	212	21420	0.26	ng/ul	0.00
Target Compounds						
5) Naphthalene	9.62	128	2037	0.04	ng/ul#	85
13) Pentachlorophenol	15.88	266	147	0.05	ng/ul#	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111015\
Data File : BE091111.D
Acq On : 11 Nov 2015 00:59
Operator : UM/NP
Sample : G4308-17
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
D9KR3

Manual Integrations
APPROVED

Mmdadoda
11/13/2015 6:29:45 PM

Quant Time: Nov 13 15:12:12 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 13 13:56:00 2015
Response via : Initial Calibration

