

Data Path : Z:\HPCHEM1\BNA E\Data\BE111015\
 Data File : BE091096.D
 Acq On : 10 Nov 2015 9:28
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.490

Quant Time: Nov 13 14:04:56 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 13 13:56:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.243	15.6	70	0.00
3	1,4-Dioxane	0.325	0.272	16.3	73	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
5	Naphthalene	0.986	0.948	3.9	96	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.541	0.9	101	0.00
7	2-Methylnaphthalene	0.643	0.634	1.4	100	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
9	Acenaphthylene	1.998	1.967	1.6	101	0.00
10 C	Acenaphthene	1.368	1.299	5.0	98	0.00
11	Fluorene	1.757	1.572	10.5	88	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
13	Pentachlorophenol	0.038	0.028	26.3#	78	0.00
14	Phenanthrene	4.619	4.341	6.0	79	0.00
15	Anthracene	4.200	4.112	2.1	86	0.00
16 SURR	Fluoranthene-d10	1.127	1.000	11.3	81	0.00
17 C	Fluoranthene	5.406	4.827	10.7	80	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	80	0.00
19	Pyrene	4.544	4.536	0.2	79	0.00
20	Benzo(a)anthracene	1.027	0.996	3.0	80	0.00
21	Chrysene	1.183	1.123	5.1	75	0.00
22 I	Perylene-d12	1.000	1.000	0.0	88	0.00
23	Benzo(b)fluoranthene	1.249	1.121	10.2	77	0.00
24	Benzo(k)fluoranthene	1.281	1.178	8.0	77	0.00
25 C	Benzo(a)pyrene	1.150	1.058	8.0	79	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.361	10.0	76	0.00
27	Dibenzo(a,h)anthracene	1.155	1.065	7.8	80	0.00
28	Benzo(q,h,i)perylene	5.083	4.675	8.0	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0