

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111015\
 Data File : BE091098.D
 Acq On : 10 Nov 2015 17:04
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.491

Quant Time: Nov 13 14:17:32 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 13 13:56:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.242	16.0	72	0.00
3	1,4-Dioxane	0.325	0.274	15.7	75	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
5	Naphthalene	0.986	0.944	4.3	97	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.537	1.6	103	0.00
7	2-Methylnaphthalene	0.643	0.612	4.8	98	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
9	Acenaphthylene	1.998	1.920	3.9	101	0.00
10 C	Acenaphthene	1.368	1.317	3.7	102	0.00
11	Fluorene	1.757	1.573	10.5	90	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
13	Pentachlorophenol	0.038	0.026	31.6#	73	0.00
14	Phenanthrene	4.619	4.410	4.5	81	0.00
15	Anthracene	4.200	4.174	0.6	88	0.00
16 SURR	Fluoranthene-d10	1.127	0.996	11.6	81	0.00
17 C	Fluoranthene	5.406	4.837	10.5	80	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
19	Pyrene	4.544	4.449	2.1	78	0.00
20	Benzo(a)anthracene	1.027	0.975	5.1	79	0.00
21	Chrysene	1.183	1.110	6.2	75	0.00
22 I	Perylene-d12	1.000	1.000	0.0	85	0.00
23	Benzo(b)fluoranthene	1.249	1.167	6.6	78	0.00
24	Benzo(k)fluoranthene	1.281	1.184	7.6	75	0.00
25 C	Benzo(a)pyrene	1.150	1.110	3.5	81	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.464	7.9	76	0.00
27	Dibenzo(a,h)anthracene	1.155	1.087	5.9	80	0.00
28	Benzo(q,h,i)perylene	5.083	4.729	7.0	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0