

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111015\
 Data File : BE091113.D
 Acq On : 11 Nov 2015 3:25
 Operator : UM/NP
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.492

Quant Time: Nov 13 15:17:30 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 13 13:56:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.239	17.0	67	0.00
3	1,4-Dioxane	0.325	0.269	17.2	70	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
5	Naphthalene	0.986	0.941	4.6	94	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.533	2.4	98	0.00
7	2-Methylnaphthalene	0.643	0.619	3.7	96	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
9	Acenaphthylene	1.998	1.950	2.4	97	0.00
10 C	Acenaphthene	1.368	1.319	3.6	96	0.00
11	Fluorene	1.757	1.584	9.8	86	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
13	Pentachlorophenol	0.038	0.031	18.4	86	0.00
14	Phenanthrene	4.619	4.304	6.8	77	0.00
15	Anthracene	4.200	4.033	4.0	83	0.00
16 SURR	Fluoranthene-d10	1.127	0.960	14.8	76	0.00
17 C	Fluoranthene	5.406	4.655	13.9	75	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	75	0.00
19	Pyrene	4.544	4.586	-0.9	75	0.00
20	Benzo(a)anthracene	1.027	0.981	4.5	74	0.00
21	Chrysene	1.183	1.125	4.9	70	0.00
22 I	Perylene-d12	1.000	1.000	0.0	79	0.00
23	Benzo(b)fluoranthene	1.249	1.162	7.0	72	0.00
24	Benzo(k)fluoranthene	1.281	1.190	7.1	70	0.00
25 C	Benzo(a)pyrene	1.150	1.101	4.3	75	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.524	6.7	72	0.00
27	Dibenzo(a,h)anthracene	1.155	1.075	6.9	73	0.00
28	Benzo(q,h,i)perylene	5.083	4.712	7.3	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0