

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE111017\
 Data File : BE094798.D
 Acq On : 9 Nov 2017 11:05
 Operator : SJ/JU
 Sample : I5976-03
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Nov 09 12:49:51 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE102517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 07 01:37:48 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	1386	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.69	136	8232	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.51	164	4769	0.40	ng/ul	0.00
10) Phenanthrene-d10	0.00	188	0	0.00	ng/ul	-17.28
16) Chrysene-d12	21.44	240	9469	0.40	ng/ul	0.00
20) Perylene-d12	23.81	264	393368	0.40	ng/ul	-0.17
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.28	152	3670	0.27	ng/ul	-0.01
14) Fluoranthene-d10	19.28	212	8326	0.00	ng/ul	0.00
Target Compounds						
17) Pyrene	19.65	202	1528	0.052	ng/ul#	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE111017\
Data File : BE094798.D
Acq On : 9 Nov 2017 11:05
Operator : SJ/JU
Sample : I5976-03
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Nov 09 12:49:51 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE102517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 07 01:37:48 2017
Response via : Initial Calibration





#17
 Pvrene
 Concen: 0.052 ng/ul
 RT: 19.65 min Scan# 2827
 Delta R.T. -0.04 min
 Lab File: BE094798.D
 Acq: 9 Nov 2017 11:05

Instrument :
 BNA_E
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
202	100		
101	143.4	18.2	27.4#
100	728.6	15.6	23.4#

