

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111115\
 Data File : BE091128.D
 Acq On : 12 Nov 2015 1:12
 Operator : UM/NP
 Sample : G4322-12
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 A4HS3

Manual Integrations
 APPROVED

Mmdadoda
 11/16/2015 12:42:11 PM

Quant Time: Nov 13 18:03:49 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 13 13:56:00 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	3397	0.40	ng/ul	0.00
4) Naphthalene-d8	9.57	136	19197	0.40	ng/ul	0.00
8) Acenaphthene-d10	13.40	164	10705	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.12	188	26536m	0.40	ng/ul	-0.01
18) Chrysene-d12	20.28	240	33998	0.40	ng/ul	0.00
22) Perylene-d12	22.19	264	29887m	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.93	96	12289	5.03	ng/uL	-0.02
6) 2-Methylnaphthalene-d10	11.08	152	3780	0.14	ng/ul	0.00
16) Fluoranthene-d10	18.12	212	10718	0.14	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111115\
Data File : BE091128.D
Acq On : 12 Nov 2015 1:12
Operator : UM/NP
Sample : G4322-12
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
A4HS3

Manual Integrations
APPROVED

Mmdadoda
11/16/2015 12:42:11 PM

Quant Time: Nov 13 18:03:49 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 13 13:56:00 2015
Response via : Initial Calibration

