

Data Path : Z:\HPCHEM1\BNA E\Data\BE111115\
 Data File : BE091115.D
 Acq On : 11 Nov 2015 17:17
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.493

Quant Time: Nov 13 15:31:00 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 13 13:56:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	-0.06
2 SURR	1,4-Dioxane-d8	0.288	0.247	14.2	63	-0.21
3	1,4-Dioxane	0.325	0.277	14.8	65	-0.21
4 I	Naphthalene-d8	1.000	1.000	0.0	92	-0.04
5	Naphthalene	0.986	0.922	6.5	83	-0.04
6 SURR	2-Methylnaphthalene-d10	0.546	0.517	5.3	87	-0.04
7	2-Methylnaphthalene	0.643	0.598	7.0	84	-0.04
8 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
9	Acenaphthylene	1.998	1.970	1.4	83	0.00
10 C	Acenaphthene	1.368	1.370	-0.1	85	0.00
11	Fluorene	1.757	1.666	5.2	77	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.03
13	Pentachlorophenol	0.038	0.024	36.8#	58	0.01
14	Phenanthrene	4.619	4.367	5.5	69	0.03
15	Anthracene	4.200	3.959	5.7	72	0.01
16 SURR	Fluoranthene-d10	1.127	0.962	14.6	67	0.03
17 C	Fluoranthene	5.406	4.744	12.2	68	0.03
18 I	Chrysene-d12	1.000	1.000	0.0	69	0.04
19	Pyrene	4.544	4.293	5.5	65	0.03
20	Benzo(a)anthracene	1.027	0.951	7.4	65	0.04
21	Chrysene	1.183	1.134	4.1	65	0.04
22 I	Perylene-d12	1.000	1.000	0.0	72	0.05
23	Benzo(b)fluoranthene	1.249	1.203	3.7	68	0.05
24	Benzo(k)fluoranthene	1.281	1.191	7.0	64	0.05
25 C	Benzo(a)pyrene	1.150	1.105	3.9	68	0.05
26	Indeno(1,2,3-cd)pyrene	4.847	4.609	4.9	66	0.09
27	Dibenzo(a,h)anthracene	1.155	1.086	6.0	67	0.09
28	Benzo(q,h,i)perylene	5.083	4.775	6.1	67	0.09

(#) = Out of Range

SPCC's out = 0 CCC's out = 0