

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111115\
 Data File : BE091138.D
 Acq On : 12 Nov 2015 10:51
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.495

Quant Time: Nov 13 16:36:21 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 13 13:56:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	-0.06
2 SURR	1,4-Dioxane-d8	0.288	0.245	14.9	62	-0.22
3	1,4-Dioxane	0.325	0.283	12.9	67	-0.21
4 I	Naphthalene-d8	1.000	1.000	0.0	91	-0.04
5	Naphthalene	0.986	0.950	3.7	85	-0.04
6 SURR	2-Methylnaphthalene-d10	0.546	0.541	0.9	89	-0.05
7	2-Methylnaphthalene	0.643	0.642	0.2	89	-0.05
8 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
9	Acenaphthylene	1.998	1.969	1.5	91	0.00
10 C	Acenaphthene	1.368	1.340	2.0	91	0.00
11	Fluorene	1.757	1.611	8.3	81	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	77	0.02
13	Pentachlorophenol	0.038	0.054	-42.1#	136	0.01
14	Phenanthrene	4.619	4.382	5.1	72	0.02
15	Anthracene	4.200	4.211	-0.3	80	0.01
16 SURR	Fluoranthene-d10	1.127	1.044	7.4	76	0.03
17 C	Fluoranthene	5.406	4.990	7.7	74	0.03
18 I	Chrysene-d12	1.000	1.000	0.0	78	0.04
19	Pyrene	4.544	4.441	2.3	76	0.03
20	Benzo(a)anthracene	1.027	1.059	-3.1	83	0.04
21	Chrysene	1.183	1.114	5.8	73	0.04
22 I	Perylene-d12	1.000	1.000	0.0	89	0.05
23	Benzo(b)fluoranthene	1.249	1.182	5.4	82	0.04
24	Benzo(k)fluoranthene	1.281	1.120	12.6	74	0.05
25 C	Benzo(a)pyrene	1.150	1.107	3.7	84	0.05
26	Indeno(1,2,3-cd)pyrene	4.847	4.366	9.9	78	0.08
27	Dibenzo(a,h)anthracene	1.155	1.094	5.3	84	0.09
28	Benzo(q,h,i)perylene	5.083	4.593	9.6	80	0.09

(#) = Out of Range

SPCC's out = 0 CCC's out = 0