

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111115\
 Data File : BE091149.D
 Acq On : 12 Nov 2015 18:18
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.496

Quant Time: Nov 13 17:26:19 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 09 05:49:45 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	-0.06
2 SURR	1,4-Dioxane-d8	0.288	0.242	16.0	69	-0.22
3	1,4-Dioxane	0.325	0.271	16.6	71	-0.21
4 I	Naphthalene-d8	1.000	1.000	0.0	99	-0.04
5	Naphthalene	0.986	0.950	3.7	92	-0.04
6 SURR	2-Methylnaphthalene-d10	0.546	0.552	-1.1	99	-0.05
7	2-Methylnaphthalene	0.643	0.648	-0.8	98	-0.05
8 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
9	Acenaphthylene	1.998	1.956	2.1	100	0.00
10 C	Acenaphthene	1.368	1.303	4.8	98	0.00
11	Fluorene	1.757	1.586	9.7	89	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.02
13	Pentachlorophenol	0.038	0.055	-44.7#	150#	0.01
14	Phenanthrene	4.619	4.405	4.6	79	0.02
15	Anthracene	4.200	4.299	-2.4	89	0.01
16 SURR	Fluoranthene-d10	1.127	1.030	8.6	82	0.03
17 C	Fluoranthene	5.406	4.924	8.9	80	0.03
18 I	Chrysene-d12	1.000	1.000	0.0	82	0.04
19	Pyrene	4.544	4.524	0.4	81	0.03
20	Benzo(a)anthracene	1.027	1.058	-3.0	87	0.04
21	Chrysene	1.183	1.111	6.1	76	0.04
22 I	Perylene-d12	1.000	1.000	0.0	93	0.05
23	Benzo(b)fluoranthene	1.249	1.164	6.8	85	0.05
24	Benzo(k)fluoranthene	1.281	1.126	12.1	78	0.05
25 C	Benzo(a)pyrene	1.150	1.108	3.7	89	0.05
26	Indeno(1,2,3-cd)pyrene	4.847	4.380	9.6	82	0.09
27	Dibenzo(a,h)anthracene	1.155	1.114	3.5	90	0.09
28	Benzo(q,h,i)perylene	5.083	4.601	9.5	84	0.09

(#) = Out of Range

SPCC's out = 0 CCC's out = 0