

Data Path : Z:\HPCHEM1\BNA E\Data\BE111315\
 Data File : BE091151.D
 Acq On : 13 Nov 2015 21:21
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.497

Quant Time: Nov 14 02:49:12 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 14 02:47:44 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.03
2 SURR	1,4-Dioxane-d8	0.288	0.234	18.7	77	0.07
3	1,4-Dioxane	0.325	0.264	18.8	81	0.06
4 I	Naphthalene-d8	1.000	1.000	0.0	102	0.02
5	Naphthalene	0.986	1.060	-7.5	107	0.02
6 SURR	2-Methylnaphthalene-d10	0.546	0.518	5.1	96	0.03
7	2-Methylnaphthalene	0.643	0.576	10.4	90	0.03
8 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.01
9	Acenaphthylene	1.998	1.800	9.9	90	0.01
10 C	Acenaphthene	1.368	1.309	4.3	97	0.01
11	Fluorene	1.757	1.586	9.7	87	0.02
12 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.01
13	Pentachlorophenol	0.038	0.052	-36.8#	155#	0.01
14	Phenanthrene	4.619	4.120	10.8	80	0.01
15	Anthracene	4.200	3.894	7.3	87	0.01
16 SURR	Fluoranthene-d10	1.127	1.104	2.0	95	0.01
17 C	Fluoranthene	5.406	5.234	3.2	92	0.01
18 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
19	Pyrene	4.544	4.053	10.8	97	0.01
20	Benzo(a)anthracene	1.027	0.966	5.9	106	0.01
21	Chrysene	1.183	1.175	0.7	108	0.01
22 I	Perylene-d12	1.000	1.000	0.0	126	0.02
23	Benzo(b)fluoranthene	1.249	1.086	13.1	107	0.01
24	Benzo(k)fluoranthene	1.281	1.254	2.1	118	0.01
25 C	Benzo(a)pyrene	1.150	1.131	1.7	122	0.01
26	Indeno(1,2,3-cd)pyrene	4.847	4.495	7.3	113	0.02
27	Dibenzo(a,h)anthracene	1.155	1.089	5.7	118	0.02
28	Benzo(q,h,i)perylene	5.083	4.747	6.6	116	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0