

Data Path : Z:\HPCHEM1\BNA E\Data\BE111315\
 Data File : BE091164.D
 Acq On : 14 Nov 2015 5:13
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.498

Quant Time: Nov 14 06:03:47 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 14 03:28:24 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	-0.01
2 SURR	1,4-Dioxane-d8	0.288	0.244	15.3	77	0.01
3	1,4-Dioxane	0.325	0.254	21.8#	74	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
5	Naphthalene	0.986	0.957	2.9	96	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.549	-0.5	102	0.00
7	2-Methylnaphthalene	0.643	0.624	3.0	98	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
9	Acenaphthylene	1.998	1.843	7.8	103	0.00
10 C	Acenaphthene	1.368	1.300	5.0	106	0.00
11	Fluorene	1.757	1.593	9.3	97	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
13	Pentachlorophenol	0.038	0.052	-36.8#	165#	0.00
14	Phenanthrene	4.619	4.311	6.7	88	0.00
15	Anthracene	4.200	4.236	-0.9	100	0.00
16 SURR	Fluoranthene-d10	1.127	1.111	1.4	101	0.00
17 C	Fluoranthene	5.406	5.306	1.8	98	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
19	Pyrene	4.544	4.411	2.9	101	0.00
20	Benzo(a)anthracene	1.027	1.019	0.8	107	0.00
21	Chrysene	1.183	1.157	2.2	102	0.00
22 I	Perylene-d12	1.000	1.000	0.0	115	0.00
23	Benzo(b)fluoranthene	1.249	1.161	7.0	105	0.00
24	Benzo(k)fluoranthene	1.281	1.204	6.0	104	0.00
25 C	Benzo(a)pyrene	1.150	1.158	-0.7	114	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.582	5.5	106	0.01
27	Dibenzo(a,h)anthracene	1.155	1.138	1.5	113	0.00
28	Benzo(q,h,i)perylene	5.083	4.806	5.4	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0