

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111315\
 Data File : BE091153.D
 Acq On : 13 Nov 2015 22:33
 Operator : UM/NP
 Sample : G4342-01
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 D9KN7

Manual Integrations
 APPROVED

mmdadoda
 11/16/2015 3:24:09 PM

Quant Time: Nov 16 10:44:26 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 14 04:34:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	4686	0.40	ng/ul	-0.07
4) Naphthalene-d8	9.57	136	22835	0.40	ng/ul	-0.04
8) Acenaphthene-d10	13.38	164	13870	0.40	ng/ul	-0.02
12) Phenanthrene-d10	16.10	188	33323m	0.40	ng/ul	-0.01
18) Chrysene-d12	20.25	240	49184	0.40	ng/ul	0.00
22) Perylene-d12	22.13	264	42607m	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	3.08	96	23595	7.00	ng/uL	-0.07
6) 2-Methylnaphthalene-d10	11.10	152	6452	0.21	ng/ul	-0.02
16) Fluoranthene-d10	18.09	212	21124	0.23	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111315\
Data File : BE091153.D
Acq On : 13 Nov 2015 22:33
Operator : UM/NP
Sample : G4342-01
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
D9KN7

Manual Integrations
APPROVED

mmdadoda
11/16/2015 3:24:09 PM

Quant Time: Nov 16 10:44:26 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 14 04:34:12 2015
Response via : Initial Calibration

