

Data Path : Z:\HPCHEM1\BNA_E\Data\BE111417\
 Data File : BE094807.D
 Acq On : 14 Nov 2017 14:27
 Operator : SJ/JU
 Sample : SSTD3.289
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD3.289

Quant Time: Nov 14 16:27:24 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE111417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 14 14:29:35 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	856	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.68	136	4791	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.51	164	2877	0.40	ng/ul	-0.01
10) Phenanthrene-d10	17.26	188	7325	0.40	ng/ul	-0.02
16) Chrysene-d12	21.43	240	8354	0.40	ng/ul	0.00
20) Perylene-d12	23.98	264	7773	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.27	152	24782	3.20	ng/ul	-0.03
14) Fluoranthene-d10	19.28	212	79032	3.14	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.73	128	35906	3.147	ng/ul#	91
5) 2-Methylnaphthalene	12.34	142	26133	3.293	ng/ul	98
7) Acenaphthylene	14.24	152	43039	3.267	ng/ul	95
8) Acenaphthene	14.57	153	29824	3.216	ng/ul	97
9) Fluorene	15.56	166	37069	3.296	ng/ul#	93
11) Pentachlorophenol	16.98	266	1208	3.110	ng/ul	95
12) Phenanthrene	17.29	178	63209	3.213	ng/ul#	91
13) Anthracene	17.38	178	65136	3.249	ng/ul#	88
15) Fluoranthene	19.31	202	80752	3.128	ng/ul	84
17) Pyrene	19.68	202	83946	3.211	ng/ul#	79
18) Benzo(a)anthracene	21.41	228	81437	3.206	ng/ul#	87
19) Chrysene	21.46	228	78016	3.077	ng/ul	98
21) Benzo(b)fluoranthene	23.18	252	77441	3.344	ng/ul	82
22) Benzo(k)fluoranthene	23.23	252	74436	3.022	ng/ul#	84
23) Benzo(a)pyrene	23.86	252	73474	3.191	ng/ul#	77
24) Indeno(1,2,3-cd)pyrene	26.66	276	80802	3.219	ng/ul#	81
25) Dibenzo(a,h)anthracene	26.66	278	67755	3.203	ng/ul#	87
26) Benzo(g,h,i)perylene	27.52	276	66769	3.175	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\Data\BE111417\
Data File : BE094807.D
Acq On : 14 Nov 2017 14:27
Operator : SJ/JU
Sample : SSTD3.289
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD3.289

Quant Time: Nov 14 16:27:24 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE111417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 14 14:29:35 2017
Response via : Initial Calibration

