

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111515\
 Data File : BE091166.D
 Acq On : 15 Nov 2015 10:45
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.499

Quant Time: Nov 16 04:50:14 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 14 03:28:24 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.260	9.7	87	0.00
3	1,4-Dioxane	0.325	0.285	12.3	88	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
5	Naphthalene	0.986	0.948	3.9	102	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.541	0.9	108	0.00
7	2-Methylnaphthalene	0.643	0.603	6.2	101	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
9	Acenaphthylene	1.998	1.860	6.9	105	0.00
10 C	Acenaphthene	1.368	1.311	4.2	108	0.00
11	Fluorene	1.757	1.637	6.8	100	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
13	Pentachlorophenol	0.038	0.036	5.3	117	0.00
14	Phenanthrene	4.619	4.429	4.1	94	0.00
15	Anthracene	4.200	4.192	0.2	103	0.00
16 SURR	Fluoranthene-d10	1.127	1.142	-1.3	108	0.00
17 C	Fluoranthene	5.406	5.504	-1.8	106	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
19	Pyrene	4.544	4.343	4.4	106	0.00
20	Benzo(a)anthracene	1.027	0.970	5.6	108	0.00
21	Chrysene	1.183	1.162	1.8	108	0.00
22 I	Perylene-d12	1.000	1.000	0.0	119	0.00
23	Benzo(b)fluoranthene	1.249	1.179	5.6	110	0.00
24	Benzo(k)fluoranthene	1.281	1.232	3.8	109	0.00
25 C	Benzo(a)pyrene	1.150	1.180	-2.6	120	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.594	5.2	109	0.00
27	Dibenzo(a,h)anthracene	1.155	1.130	2.2	116	0.00
28	Benzo(g,h,i)perylene	5.083	4.895	3.7	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0