

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111515\
 Data File : BE091174.D
 Acq On : 15 Nov 2015 15:52
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.467

Quant Time: Nov 16 04:57:05 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 16 04:54:02 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	-0.01
2 SURR	1,4-Dioxane-d8	0.288	0.243	15.6	86	-0.02
3	1,4-Dioxane	0.325	0.259	20.3#	85	-0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
5	Naphthalene	0.986	0.947	4.0	112	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.537	1.6	118	0.00
7	2-Methylnaphthalene	0.643	0.608	5.4	112	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	129	0.00
9	Acenaphthylene	1.998	1.847	7.6	116	0.00
10 C	Acenaphthene	1.368	1.280	6.4	118	0.00
11	Fluorene	1.757	1.611	8.3	110	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
13	Pentachlorophenol	0.038	0.043	-13.2	153#	-0.01
14	Phenanthrene	4.619	4.402	4.7	102	0.00
15	Anthracene	4.200	4.059	3.4	108	0.00
16 SURR	Fluoranthene-d10	1.127	1.092	3.1	112	0.00
17 C	Fluoranthene	5.406	5.223	3.4	109	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
19	Pyrene	4.544	4.392	3.3	109	0.00
20	Benzo(a)anthracene	1.027	0.978	4.8	111	0.00
21	Chrysene	1.183	1.146	3.1	109	0.00
22 I	Perylene-d12	1.000	1.000	0.0	120	0.00
23	Benzo(b)fluoranthene	1.249	1.194	4.4	112	0.00
24	Benzo(k)fluoranthene	1.281	1.200	6.3	108	0.00
25 C	Benzo(a)pyrene	1.150	1.198	-4.2	123	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.517	6.8	109	0.00
27	Dibenzo(a,h)anthracene	1.155	1.138	1.5	118	0.00
28	Benzo(g,h,i)perylene	5.083	4.838	4.8	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0