

Data Path : Z:\HPCHEM1\BNA_E\Data\BE111517\
 Data File : BE094810.D
 Acq On : 15 Nov 2017 9:48
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.491

Quant Time: Nov 16 00:56:12 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE111417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 14 16:33:16 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
3	Naphthalene	0.953	0.889	6.7	107	0.00
4 SURR	2-Methylnaphthalene-d10	0.646	0.586	9.3	105	0.00
5	2-Methylnaphthalene	0.663	0.610	8.0	106	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
7	Acenaphthylene	1.832	1.688	7.9	106	0.00
8 C	Acenaphthene	1.289	1.216	5.7	108	0.00
9	Fluorene	1.564	1.433	8.4	104	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
11	Pentachlorophenol	0.021	0.016	23.8#	86	0.00
12	Phenanthrene	1.074	0.966	10.1	106	0.00
13	Anthracene	1.095	0.969	11.5	103	0.00
14 SURR	Fluoranthene-d10	1.374	1.279	6.9	107	0.00
15 C	Fluoranthene	1.410	1.315	6.7	107	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	125	0.00
17	Pyrene	1.252	1.102	12.0	109	0.00
18	Benzo(a)anthracene	1.216	1.107	9.0	111	0.00
19	Chrysene	1.214	1.098	9.6	111	0.00
20 I	Perylene-d12	1.000	1.000	0.0	129	0.00
21	Benzo(b)fluoranthene	1.192	1.049	12.0	107	0.00
22	Benzo(k)fluoranthene	1.267	1.163	8.2	119	0.00
23 C	Benzo(a)pyrene	1.185	1.061	10.5	112	0.00
24	Indeno(1,2,3-cd)pyrene	1.292	1.158	10.4	112	0.00
25	Dibenzo(a,h)anthracene	1.089	0.974	10.6	111	0.00
26	Benzo(g,h,i)perylene	1.082	0.968	10.5	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0