

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111616\
 Data File : BE092542.D
 Acq On : 16 Nov 2016 20:20
 Operator : UM/SJ
 Sample : H5636-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 CPS073041

Quant Time: Nov 17 05:00:26 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	7671	5.00	ng	-0.02
7) Naphthalene-d8	10.93	136	34846	5.00	ng	0.00
12) Acenaphthene-d10	14.80	164	23977	5.00	ng	0.00
18) Phenanthrene-d10	17.56	188	50508	5.00	ng	0.00
24) Chrysene-d12	21.72	240	78331	5.00	ng	0.00
31) Perylene-d12	24.08	264	69231	5.00	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.65	112	8307	6.41	ng	-0.02
5) Phenol-d6	7.33	99	6561	3.10	ng	-0.02
8) Nitrobenzene-d5	9.32	82	12318	5.44	ng	-0.05
13) 2,4,6-Tribromophenol	16.29	330	7243	13.47	ng	-0.01
14) 2-Fluorobiphenyl	13.42	172	31723	5.43	ng	-0.01
26) Terphenyl-d14	20.16	244	55061	6.15	ng	-0.02
Target Compounds						
20) Pentachlorophenol	17.23	266	398	1.08	ng	Qvalue 97
29) Bis(2-ethylhexyl)phthalate	21.61	149	545	0.23	ng	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111616\
Data File : BE092542.D
Acq On : 16 Nov 2016 20:20
Operator : UM/SJ
Sample : H5636-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
CPS073041

Quant Time: Nov 17 05:00:26 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 16 15:43:18 2016
Response via : Initial Calibration

