

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111616\
 Data File : BE092536.D
 Acq On : 16 Nov 2016 15:22
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE111616

Quant Time: Nov 16 16:07:56 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.681	0.729	-7.0	115	0.00
3	n-Nitrosodimethylamine	0.654	0.559	14.5	103	0.00
4 S	2-Fluorophenol	0.844	0.839	0.6	103	0.00
5 S	Phenol-d6	1.104	1.038	6.0	105	0.00
6	bis(2-Chloroethyl)ether	1.341	1.195	10.9	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.269	0.252	6.3	97	0.00
9	Naphthalene	1.032	1.020	1.2	100	0.00
10	Hexachlorobutadiene	0.157	0.155	1.3	99	0.00
11	2-Methylnaphthalene	0.658	0.617	6.2	100	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
13 S	2,4,6-Tribromophenol	0.112	0.098	12.5	98	0.00
14 S	2-Fluorobiphenyl	1.219	1.155	5.3	100	0.00
15	Acenaphthylene	6.962	6.303	9.5	100	0.00
16	Acenaphthene	1.136	1.085	4.5	101	0.00
17	Fluorene	1.408	1.315	6.6	100	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
19	Hexachlorobenzene	0.197	0.183	7.1	101	0.00
20	Pentachlorophenol	0.036	0.033	8.3	113	0.00
21	Phenanthrene	1.119	1.089	2.7	98	0.00
22	Anthracene	1.053	1.015	3.6	100	0.00
23	Fluoranthene	5.219	4.950	5.2	99	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
25	Pyrene	3.975	3.567	10.3	98	0.00
26 S	Terphenyl-d14	0.572	0.505	11.7	97	0.00
27	Benzo(a)anthracene	4.692	4.135	11.9	106	0.00
28	Chrysene	4.250	4.536	-6.7	104	-0.02
29	Bis(2-ethylhexyl)phthalate	0.305	0.283	7.2	113	-0.02
30	Indeno(1,2,3-cd)pyrene	4.722	4.238	10.2	96	0.00
31 I	Perylene-d12	1.000	1.000	0.0	98	-0.01
32	Benzo(b)fluoranthene	1.213	1.186	2.2	100	0.00
33	Benzo(k)fluoranthene	1.270	1.138	10.4	93	0.00
34 C	Benzo(a)pyrene	1.122	1.020	9.1	96	0.00
35	Dibenzo(a,h)anthracene	1.111	1.003	9.7	95	0.00
36	Benzo(g,h,i)perylene	4.702	4.369	7.1	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0