

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111616\
 Data File : BE092538.D
 Acq On : 16 Nov 2016 17:51
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 17 14:09:20 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	-0.02
2	1,4-Dioxane	0.681	0.771	-13.2	120	0.00
3	n-Nitrosodimethylamine	0.654	0.629	3.8	116	0.00
4 S	2-Fluorophenol	0.844	0.840	0.5	102	0.00
5 S	Phenol-d6	1.104	1.034	6.3	104	0.00
6	bis(2-Chloroethyl)ether	1.341	1.148	14.4	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.269	0.263	2.2	102	0.00
9	Naphthalene	1.032	1.043	-1.1	103	0.00
10	Hexachlorobutadiene	0.157	0.163	-3.8	105	0.00
11	2-Methylnaphthalene	0.658	0.642	2.4	105	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
13 S	2,4,6-Tribromophenol	0.112	0.098	12.5	97	0.00
14 S	2-Fluorobiphenyl	1.219	1.209	0.8	104	0.00
15	Acenaphthylene	6.962	6.623	4.9	104	0.00
16	Acenaphthene	1.136	1.115	1.8	103	0.00
17	Fluorene	1.408	1.388	1.4	105	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
19	Hexachlorobenzene	0.197	0.198	-0.5	110	0.00
20	Pentachlorophenol	0.036	0.034	5.6	119	0.00
21	Phenanthrene	1.119	1.140	-1.9	104	0.00
22	Anthracene	1.053	1.052	0.1	105	0.00
23	Fluoranthene	5.219	5.246	-0.5	105	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
25	Pyrene	3.975	3.700	6.9	103	0.00
26 S	Terphenyl-d14	0.572	0.533	6.8	105	0.00
27	Benzo(a)anthracene	4.692	4.194	10.6	110	0.00
28	Chrysene	4.250	4.313	-1.5	102	-0.02
29	Bis(2-ethylhexyl)phthalate	0.305	0.269	11.8	109	-0.02
30	Indeno(1,2,3-cd)pyrene	4.722	4.393	7.0	101	0.00
31 I	Perylene-d12	1.000	1.000	0.0	101	-0.01
32	Benzo(b)fluoranthene	1.213	1.224	-0.9	106	0.00
33	Benzo(k)fluoranthene	1.270	1.168	8.0	99	0.00
34 C	Benzo(a)pyrene	1.122	1.040	7.3	102	0.00
35	Dibenzo(a,h)anthracene	1.111	1.024	7.8	100	-0.02
36	Benzo(g,h,i)perylene	4.702	4.468	5.0	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0