

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111616\
 Data File : BE092551.D
 Acq On : 17 Nov 2016 1:55
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 17 05:10:47 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	-0.02
2	1,4-Dioxane	0.681	0.850	-24.8	131	0.00
3	n-Nitrosodimethylamine	0.654	0.673	-2.9	122	-0.01
4 S	2-Fluorophenol	0.844	1.017	-20.5	122	0.00
5 S	Phenol-d6	1.104	1.308	-18.5	130	0.00
6	bis(2-Chloroethyl)ether	1.341	1.207	10.0	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.269	0.294	-9.3	110	-0.02
9	Naphthalene	1.032	1.037	-0.5	100	-0.02
10	Hexachlorobutadiene	0.157	0.168	-7.0	105	0.00
11	2-Methylnaphthalene	0.658	0.668	-1.5	106	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
13 S	2,4,6-Tribromophenol	0.112	0.124	-10.7	122	-0.01
14 S	2-Fluorobiphenyl	1.219	1.222	-0.2	103	0.00
15	Acenaphthylene	6.962	6.935	0.4	108	0.00
16	Acenaphthene	1.136	1.142	-0.5	104	0.00
17	Fluorene	1.408	1.439	-2.2	107	-0.01
18 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
19	Hexachlorobenzene	0.197	0.196	0.5	109	0.00
20	Pentachlorophenol	0.036	0.047	-30.6#	163#	0.00
21	Phenanthrene	1.119	1.136	-1.5	104	-0.02
22	Anthracene	1.053	1.100	-4.5	109	0.00
23	Fluoranthene	5.219	5.535	-6.1	111	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
25	Pyrene	3.975	3.850	3.1	112	0.00
26 S	Terphenyl-d14	0.572	0.541	5.4	111	-0.02
27	Benzo(a)anthracene	4.692	4.386	6.5	120	0.00
28	Chrysene	4.250	4.392	-3.3	108	-0.02
29	Bis(2-ethylhexyl)phthalate	0.305	0.393	-28.9#	167#	-0.02
30	Indeno(1,2,3-cd)pyrene	4.722	4.569	3.2	110	0.00
31 I	Perylene-d12	1.000	1.000	0.0	105	-0.01
32	Benzo(b)fluoranthene	1.213	1.149	5.3	103	-0.01
33	Benzo(k)fluoranthene	1.270	1.309	-3.1	115	-0.01
34 C	Benzo(a)pyrene	1.122	1.117	0.4	113	0.00
35	Dibenzo(a,h)anthracene	1.111	1.097	1.3	111	-0.02
36	Benzo(g,h,i)perylene	4.702	4.638	1.4	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0