

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111715\
 Data File : BE091207.D
 Acq On : 17 Nov 2015 16:54
 Operator : UM/NP
 Sample : G4309-25
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 D9KN5

Manual Integrations
 APPROVED

Mmdadoda
 11/19/2015 6:41:40 PM

Quant Time: Nov 18 04:30:55 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 18 04:23:52 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	3289	0.40	ng/ul	-0.03
4) Naphthalene-d8	9.55	136	17889	0.40	ng/ul	-0.02
8) Acenaphthene-d10	13.36	164	11336	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.09	188	31095m	0.40	ng/ul	0.00
18) Chrysene-d12	20.23	240	50424	0.40	ng/ul	0.00
22) Perylene-d12	22.11	264	59568m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) 1,4-Dioxane-d8	2.98	96	18474	7.81	ng/uL	-0.03
6) 2-Methylnaphthalene-d10	11.07	152	6234	0.26	ng/ul	0.00
16) Fluoranthene-d10	18.08	212	20797	0.24	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111715\
Data File : BE091207.D
Acq On : 17 Nov 2015 16:54
Operator : UM/NP
Sample : G4309-25
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
D9KN5

Manual Integrations
APPROVED

Mmdadoda
11/19/2015 6:41:40 PM

Quant Time: Nov 18 04:30:55 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 18 04:23:52 2015
Response via : Initial Calibration

