

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111715\
 Data File : BE091205.D
 Acq On : 17 Nov 2015 15:30
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.471

Quant Time: Nov 18 03:30:19 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 18 03:27:44 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.241	16.3	88	0.02
3	1,4-Dioxane	0.325	0.269	17.2	92	0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
5	Naphthalene	0.986	0.970	1.6	117	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.516	5.5	115	0.00
7	2-Methylnaphthalene	0.643	0.603	6.2	113	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
9	Acenaphthylene	1.998	1.941	2.9	118	0.00
10 C	Acenaphthene	1.368	1.347	1.5	120	0.00
11	Fluorene	1.757	1.723	1.9	114	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
13	Pentachlorophenol	0.038	0.056	-47.4#	206#	0.00
14	Phenanthrene	4.619	4.415	4.4	106	0.00
15	Anthracene	4.200	4.221	-0.5	117	0.00
16 SURR	Fluoranthene-d10	1.127	1.123	0.4	119	0.00
17 C	Fluoranthene	5.406	5.488	-1.5	119	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
19	Pyrene	4.544	4.272	6.0	116	0.00
20	Benzo(a)anthracene	1.027	0.998	2.8	124	0.00
21	Chrysene	1.183	1.151	2.7	120	0.00
22 I	Perylene-d12	1.000	1.000	0.0	133	0.00
23	Benzo(b)fluoranthene	1.249	1.177	5.8	123	0.00
24	Benzo(k)fluoranthene	1.281	1.223	4.5	122	0.00
25 C	Benzo(a)pyrene	1.150	1.211	-5.3	138	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.604	5.0	123	0.00
27	Dibenzo(a,h)anthracene	1.155	1.135	1.7	130	0.00
28	Benzo(g,h,i)perylene	5.083	4.806	5.4	125	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0