

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111715\
 Data File : BE091218.D
 Acq On : 18 Nov 2015 1:09
 Operator : UM/NP
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.472

Quant Time: Nov 18 04:27:32 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 18 04:23:52 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	0#	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.218	24.3#	0#	0.01
3	1,4-Dioxane	0.325	0.145	55.4#	0#	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	0#	-0.02
5	Naphthalene	0.986	14.500	-1370.6#	0#	-0.07
6 SURR	2-Methylnaphthalene-d10	0.546	3.500	-541.0#	0#	-0.03
7	2-Methylnaphthalene	0.643	3.500	-444.3#	0#	-0.02
8 I	Acenaphthene-d10	1.000	1.000	0.0	0#	-0.03
9	Acenaphthylene	1.998	2.800	-40.1#	0#	-0.06
10 C	Acenaphthene	1.368	1.800	-31.6#	0#	-0.03
11	Fluorene	1.757	2.000	-13.8	0#	-0.05
12 I	Phenanthrene-d10	1.000	1.000	0.0	0#	0.01
13	Pentachlorophenol	0.038	0.120	-215.8#	0#	0.01
14	Phenanthrene	4.619	1.520	67.1#	0#	-0.03
15	Anthracene	4.200	14.680	-249.5#	0#	0.00
16 SURR	Fluoranthene-d10	1.127	0.120	89.4#	0#	0.03
17 C	Fluoranthene	5.406	0.360	93.3#	0#	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	0#	0.02
19	Pyrene	4.544	11.556	-154.3#	0#	0.05
20	Benzo(a)anthracene	1.027	0.556	45.9#	0#	0.00
21	Chrysene	1.183	0.778	34.2#	0#	0.02
22 I	Perylene-d12	1.000	1.000	0.0	0#	0.03
23	Benzo(b)fluoranthene	1.249	2.400	-92.2#	0#	0.03
24	Benzo(k)fluoranthene	1.281	2.200	-71.7#	0#	0.03
25 C	Benzo(a)pyrene	1.150	3.400	-195.7#	0#	0.01
26	Indeno(1,2,3-cd)pyrene	4.847	5.200	-7.3	0#	0.04
27	Dibenzo(a,h)anthracene	1.155	1.400	-21.2#	0#	0.03
28	Benzo(g,h,i)perylene	5.083	9.200	-81.0#	0#	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 3