

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111716\
 Data File : BE092560.D
 Acq On : 17 Nov 2016 15:38
 Operator : UM/SJ
 Sample : H5636-08DL 10X
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C19023041DL

Quant Time: Nov 18 13:34:10 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	8054	5.00	ng	-0.02
7) Naphthalene-d8	10.93	136	38907	5.00	ng	0.00
12) Acenaphthene-d10	14.80	164	29076	5.00	ng	0.00
18) Phenanthrene-d10	17.56	188	61854	5.00	ng	0.00
24) Chrysene-d12	21.72	240	93002	5.00	ng	0.00
31) Perylene-d12	24.08	264	76420	5.00	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.66	112	655	0.48	ng	-0.02
5) Phenol-d6	7.34	99	517	0.34	ng	-0.02
8) Nitrobenzene-d5	9.34	82	1132	0.52	ng	-0.02
13) 2,4,6-Tribromophenol	16.30	330	776	1.19	ng	-0.01
14) 2-Fluorobiphenyl	13.42	172	3610	0.51	ng	-0.01
26) Terphenyl-d14	20.16	244	6360	0.60	ng	-0.02
Target Compounds						Qvalue
9) Naphthalene	10.97	128	9731	1.21	ng	99
11) 2-Methylnaphthalene	12.60	142	2498	0.49	ng	97
16) Acenaphthene	14.85	154	2147	0.33	ng	# 81
17) Fluorene	15.85	166	4488	0.55	ng	# 96
21) Phenanthrene	17.58	178	7430	0.54	ng	99
29) Bis(2-ethylhexyl)phthalate	21.61	149	166	0.15	ng	# 56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111716\
Data File : BE092560.D
Acq On : 17 Nov 2016 15:38
Operator : UM/SJ
Sample : H5636-08DL 10X
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C19023041DL

Quant Time: Nov 18 13:34:10 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE111616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 16 15:43:18 2016
Response via : Initial Calibration

