

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111815\
 Data File : BE091223.D
 Acq On : 18 Nov 2015 12:15
 Operator : UM/NP
 Sample : G4309-26
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 D9KN6

Manual Integrations
 APPROVED

Mmdadoda
 11/19/2015 6:46:28 PM

Quant Time: Nov 19 17:29:53 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 19 04:25:56 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	3430	0.40	ng/ul	-0.02
4) Naphthalene-d8	9.55	136	17365	0.40	ng/ul	-0.02
8) Acenaphthene-d10	13.36	164	9181	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.09	188	25338m	0.40	ng/ul	0.00
18) Chrysene-d12	20.23	240	48422	0.40	ng/ul	0.00
22) Perylene-d12	22.11	264	40493m	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.97	96	20608	8.35	ng/uL	-0.02
6) 2-Methylnaphthalene-d10	11.07	152	5819	0.25	ng/ul	0.00
16) Fluoranthene-d10	18.07	212	18153	0.25	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111815\
Data File : BE091223.D
Acq On : 18 Nov 2015 12:15
Operator : UM/NP
Sample : G4309-26
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
D9KN6

Manual Integrations
APPROVED

Mmdadoda
11/19/2015 6:46:28 PM

Quant Time: Nov 19 17:29:53 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 19 04:25:56 2015
Response via : Initial Calibration

