

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111815\
 Data File : BE091230.D
 Acq On : 18 Nov 2015 16:31
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.474

Quant Time: Nov 19 04:30:18 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 19 04:25:56 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.91	152	5556	0.40	ng/ul	0.00
4) Naphthalene-d8	9.57	136	27350	0.40	ng/ul	0.00
8) Acenaphthene-d10	13.37	164	12988	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.09	188	37633	0.40	ng/ul	0.00
18) Chrysene-d12	20.23	240	44578	0.40	ng/ul	0.00
22) Perylene-d12	22.11	264	36744	0.40	ng/ul	0.00

System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.99	96	6737	1.69	ng/uL	0.00
6) 2-Methylnaphthalene-d10	11.07	152	12630	0.34	ng/ul	0.00
16) Fluoranthene-d10	18.07	212	46955	0.44	ng/ul	0.00

Target Compounds				Qvalue		
3) 1,4-Dioxane	3.01	88	7645	1.69	ng/ul#	65
5) Naphthalene	9.61	128	25809	0.38	ng/ul	98
7) 2-Methylnaphthalene	11.13	142	15048	0.34	ng/ul	88
9) Acenaphthylene	13.12	152	25569	0.39	ng/ul	98
10) Acenaphthene	13.43	153	18718	0.42	ng/ul#	86
11) Fluorene	14.42	166	24926	0.44	ng/ul	91
13) Pentachlorophenol	15.86	266	1105	0.31	ng/ul	95
14) Phenanthrene	16.13	178	173318	0.40	ng/ul	98
15) Anthracene	16.21	178	155556	0.39	ng/ul	97
17) Fluoranthene	18.10	202	222073	0.44	ng/ul	88
19) Pyrene	18.49	202	188997	0.37	ng/ul#	79
20) Benzo(a)anthracene	20.20	228	45037	0.39	ng/ul	99
21) Chrysene	20.26	228	53174	0.40	ng/ul	97
23) Benzo(b)fluoranthene	21.56	252	47392	0.41	ng/ul	94
24) Benzo(k)fluoranthene	21.58	252	45661	0.39	ng/ul	94
25) Benzo(a)pyrene	22.01	252	45733	0.43	ng/ul	92
26) Indeno(1,2,3-cd)pyrene	23.70	276	170576	0.38	ng/ul#	91
27) Dibenzo(a,h)anthracene	23.66	278	41714	0.39	ng/ul#	71
28) Benzo(g,h,i)perylene	24.26	276	176598	0.38	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111815\
Data File : BE091230.D
Acq On : 18 Nov 2015 16:31
Operator : UM/NP
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.474

Quant Time: Nov 19 04:30:18 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 19 04:25:56 2015
Response via : Initial Calibration

