

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111815\
 Data File : BE091238.D
 Acq On : 18 Nov 2015 21:24
 Operator : UM/NP
 Sample : G4310-02
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 D9KL7

Manual Integrations
 APPROVED

Mmdadoda
 11/19/2015 7:01:06 PM

Quant Time: Nov 19 18:34:34 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 19 06:02:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	4038	0.40	ng/ul	-0.03
4) Naphthalene-d8	9.55	136	21052	0.40	ng/ul	-0.02
8) Acenaphthene-d10	13.36	164	13498	0.40	ng/ul	0.00
12) Phenanthrene-d10	16.09	188	38541m	0.40	ng/ul	0.00
18) Chrysene-d12	20.23	240	69109	0.40	ng/ul	0.00
22) Perylene-d12	22.11	264	58933m	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.96	96	22263	7.66	ng/uL	-0.03
6) 2-Methylnaphthalene-d10	11.07	152	6609	0.23	ng/ul	0.00
16) Fluoranthene-d10	18.07	212	22445	0.21	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE111815\
Data File : BE091238.D
Acq On : 18 Nov 2015 21:24
Operator : UM/NP
Sample : G4310-02
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
D9KL7

Manual Integrations
APPROVED

Mmdadoda
11/19/2015 7:01:06 PM

Quant Time: Nov 19 18:34:34 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 19 06:02:12 2015
Response via : Initial Calibration

