

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111815\
 Data File : BE091220.D
 Acq On : 18 Nov 2015 10:27
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.473

Quant Time: Nov 19 04:24:35 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 18 04:23:52 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	134	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.248	13.9	108	0.00
3	1,4-Dioxane	0.325	0.279	14.2	112	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	139	0.00
5	Naphthalene	0.986	0.944	4.3	129	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.456	16.5	115	0.00
7	2-Methylnaphthalene	0.643	0.544	15.4	116	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
9	Acenaphthylene	1.998	1.990	0.4	114	0.00
10 C	Acenaphthene	1.368	1.454	-6.3	123	0.00
11	Fluorene	1.757	1.954	-11.2	122	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
13	Pentachlorophenol	0.038	0.025	34.2#	95	0.00
14	Phenanthrene	4.619	4.692	-1.6	117	0.00
15	Anthracene	4.200	4.059	3.4	116	0.00
16 SURR	Fluoranthene-d10	1.127	1.312	-16.4	145	0.00
17 C	Fluoranthene	5.406	6.162	-14.0	139	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	134	0.00
19	Pyrene	4.544	4.109	9.6	120	0.00
20	Benzo(a)anthracene	1.027	0.977	4.9	131	0.00
21	Chrysene	1.183	1.194	-0.9	134	0.00
22 I	Perylene-d12	1.000	1.000	0.0	128	0.00
23	Benzo(b)fluoranthene	1.249	1.300	-4.1	130	0.00
24	Benzo(k)fluoranthene	1.281	1.276	0.4	121	0.00
25 C	Benzo(a)pyrene	1.150	1.267	-10.2	138	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.759	1.8	121	0.00
27	Dibenzo(a,h)anthracene	1.155	1.154	0.1	127	0.00
28	Benzo(g,h,i)perylene	5.083	4.872	4.2	121	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0