

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111815\
 Data File : BE091230.D
 Acq On : 18 Nov 2015 16:31
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.474

Quant Time: Nov 19 04:30:18 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 19 04:25:56 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	139	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.243	15.6	110	0.00
3	1,4-Dioxane	0.325	0.275	15.4	115	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	146	0.00
5	Naphthalene	0.986	0.944	4.3	136	0.00
6 SURR	2-Methylnaphthalene-d10	0.546	0.462	15.4	123	0.00
7	2-Methylnaphthalene	0.643	0.550	14.5	123	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	130	0.00
9	Acenaphthylene	1.998	1.969	1.5	124	0.00
10 C	Acenaphthene	1.368	1.441	-5.3	134	0.00
11	Fluorene	1.757	1.919	-9.2	132	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
13	Pentachlorophenol	0.038	0.029	23.7#	121	0.00
14	Phenanthrene	4.619	4.605	0.3	123	0.00
15	Anthracene	4.200	4.133	1.6	127	0.00
16 SURR	Fluoranthene-d10	1.127	1.248	-10.7	147	0.00
17 C	Fluoranthene	5.406	5.901	-9.2	142	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	136	0.00
19	Pyrene	4.544	4.240	6.7	126	0.00
20	Benzo(a)anthracene	1.027	1.010	1.7	138	0.00
21	Chrysene	1.183	1.193	-0.8	136	0.00
22 I	Perylene-d12	1.000	1.000	0.0	133	0.00
23	Benzo(b)fluoranthene	1.249	1.290	-3.3	134	0.00
24	Benzo(k)fluoranthene	1.281	1.243	3.0	123	0.00
25 C	Benzo(a)pyrene	1.150	1.245	-8.3	141	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.642	4.2	123	0.00
27	Dibenzo(a,h)anthracene	1.155	1.135	1.7	129	0.00
28	Benzo(g,h,i)perylene	5.083	4.806	5.4	124	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0