

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111815\
 Data File : BE091244.D
 Acq On : 19 Nov 2015 1:49
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.475

Quant Time: Nov 19 07:12:26 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 19 06:02:12 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	170#	-0.03
2 SURR	1,4-Dioxane-d8	0.288	0.247	14.2	136	-0.03
3	1,4-Dioxane	0.325	0.264	18.8	135	-0.04
4 I	Naphthalene-d8	1.000	1.000	0.0	194#	-0.02
5	Naphthalene	0.986	0.964	2.2	185#	-0.02
6 SURR	2-Methylnaphthalene-d10	0.546	0.551	-0.9	195#	-0.02
7	2-Methylnaphthalene	0.643	0.650	-1.1	194#	-0.02
8 I	Acenaphthene-d10	1.000	1.000	0.0	208#	0.00
9	Acenaphthylene	1.998	1.950	2.4	197#	-0.02
10 C	Acenaphthene	1.368	1.335	2.4	199#	0.00
11	Fluorene	1.757	1.656	5.7	183#	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	172#	0.00
13	Pentachlorophenol	0.038	0.067	-76.3#	382#	0.00
14	Phenanthrene	4.619	4.479	3.0	165#	0.00
15	Anthracene	4.200	4.392	-4.6	186#	-0.01
16 SURR	Fluoranthene-d10	1.127	1.072	4.9	175#	0.00
17 C	Fluoranthene	5.406	5.203	3.8	173#	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	176#	0.00
19	Pyrene	4.544	4.683	-3.1	179#	0.00
20	Benzo(a)anthracene	1.027	1.090	-6.1	192#	0.00
21	Chrysene	1.183	1.130	4.5	166#	0.00
22 I	Perylene-d12	1.000	1.000	0.0	194#	0.00
23	Benzo(b)fluoranthene	1.249	1.155	7.5	175#	0.00
24	Benzo(k)fluoranthene	1.281	1.167	8.9	168#	0.00
25 C	Benzo(a)pyrene	1.150	1.124	2.3	186#	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.294	11.4	166#	0.00
27	Dibenzo(a,h)anthracene	1.155	1.112	3.7	185#	0.00
28	Benzo(g,h,i)perylene	5.083	4.548	10.5	171#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0