

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111816\
 Data File : BE092569.D
 Acq On : 18 Nov 2016 16:57
 Operator : UM/SJ
 Sample : PB94794BSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB94794BSD

Quant Time: Nov 18 23:23:38 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.14	152	10026	5.00	ng	-0.02
7) Naphthalene-d8	10.93	136	45697	5.00	ng	0.00
12) Acenaphthene-d10	14.80	164	29360	5.00	ng	0.00
18) Phenanthrene-d10	17.56	188	53078	5.00	ng	0.00
24) Chrysene-d12	21.72	240	68066	5.00	ng	0.00
31) Perylene-d12	24.08	264	59429	5.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.65	112	20257	11.96	ng	-0.02
5) Phenol-d6	7.31	99	29680	10.47	ng	-0.05
8) Nitrobenzene-d5	9.31	82	13601	4.59	ng	-0.05
13) 2,4,6-Tribromophenol	16.28	330	6842	10.39	ng	-0.02
14) 2-Fluorobiphenyl	13.41	172	30495	4.26	ng	-0.02
26) Terphenyl-d14	20.16	244	35205	4.52	ng	-0.02


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.39	88	4311	3.64	ng	# 52
3) n-Nitrosodimethylamine	3.73	42	6909	4.15	ng	# 97
6) bis(2-Chloroethyl)ether	7.56	93	12475	4.09	ng	# 86
9) Naphthalene	10.97	128	37023	3.93	ng	# 95
10) Hexachlorobutadiene	11.26	225	5448	3.80	ng	100
11) 2-Methylnaphthalene	12.60	142	24680	4.10	ng	93
15) Acenaphthylene	14.51	152	165471	4.05	ng	100
16) Acenaphthene	14.85	154	24306	3.64	ng	92
17) Fluorene	15.85	166	32363	3.92	ng	99
19) Hexachlorobenzene	16.87	284	9732	4.64	ng	# 65
20) Pentachlorophenol	17.21	266	7243	7.36	ng	# 85
21) Phenanthrene	17.58	178	53391	4.50	ng	# 94
22) Anthracene	17.67	178	51378	4.59	ng	99
23) Fluoranthene	19.58	202	217373	3.92	ng	98
25) Pyrene	19.95	202	230871	4.27	ng	99
27) Benzo(a)anthracene	21.70	228	505385	7.78	ng	# 90
28) Chrysene	21.70	228	506768	8.76	ng	# 71
29) Bis(2-ethylhexyl)phthalate	21.61	149	22725	5.14	ng	# 75
30) Indeno(1,2,3-cd)pyrene	26.55	276	298018	4.72	ng	98
32) Benzo(b)fluoranthene	23.35	252	59507	4.13	ng	# 97
33) Benzo(k)fluoranthene	23.39	252	56610	3.75	ng	94
34) Benzo(a)pyrene	23.96	252	57112	4.28	ng	# 94
35) Dibenzo(a,h)anthracene	26.56	278	61326	4.65	ng	96
36) Benzo(g,h,i)perylene	27.31	276	258350	4.62	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE111816\
Data File : BE092569.D
Acq On : 18 Nov 2016 16:57
Operator : UM/SJ
Sample : PB94794BSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB94794BSD

Quant Time: Nov 18 23:23:38 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE111616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 16 15:43:18 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.14 min Scan# 575

Delta R.T. -0.02 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57

Instrument :

BNA_E

ClientSampleId :

PB94794BSD


Tgt Ion:152 Resp: 10026

Ion Ratio Lower Upper

152 100

150 183.4 106.0 159.0#




115 88.2 31.2 46.8#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 3.64 ng

RT: 3.39 min Scan# 69

Delta R.T. 0.00 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57


Tgt Ion: 88 Resp: 4311

Ion Ratio Lower Upper

88 100

58 102.3 63.6 95.4#



43 91.3 28.0 42.0#




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 4.15 ng

RT: 3.73 min Scan# 99

Delta R.T. -0.05 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57

Instrument :

BNA_E

ClientSampleId :

PB94794BSD


Tgt Ion: 42 Resp: 6909

Ion Ratio Lower Upper

42 100

74 104.5 86.0 129.0




44 9.9 5.1 7.7#



Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0



#4

2-Fluorophenol

Concen: 11.96 ng

RT: 5.65 min Scan# 360

Delta R.T. -0.02 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57


Tgt Ion: 112 Resp: 20257

Ion Ratio Lower Upper

112 100

64 68.4 53.5 80.3



63 36.7 27.9 41.9




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 10.47 ng

RT: 7.31 min Scan# 539

Delta R.T. -0.05 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57

Instrument :

BNA_E

ClientSampleId :

PB94794BSD


Tgt Ion: 99 Resp: 29680

Ion Ratio Lower Upper

99 100

42 22.7 16.0 24.0




71 33.9 30.6 45.8



Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 4.09 ng

RT: 7.56 min Scan# 550

Delta R.T. -0.02 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57


Tgt Ion: 93 Resp: 12475

Ion Ratio Lower Upper

93 100

63 73.4 71.6 107.4



95 32.1 24.0 36.0




Abundance Ion 93.00 (92.70 to 93.70): BE0

Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.93 min Scan# 707

Delta R.T. -0.00 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57


Instrument :

BNA_E

ClientSampleId :

PB94794BSD

Tgt Ion:	136	Resp:	45697
Ion	Ratio	Lower	Upper
136	100		
137	11.9	8.2	12.4
54	5.3	9.6	14.4#
68	4.2	6.7	10.1#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 4.59 ng


RT: 9.31 min Scan# 626

Delta R.T. -0.05 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57



Tgt Ion:	82	Resp:	13601
Ion	Ratio	Lower	Upper
82	100		
128	32.5	39.0	58.4#
54	47.8	44.8	67.2




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 3.93 ng

RT: 10.97 min Scan# 709

Delta R.T. -0.02 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57

Instrument :

BNA_E

ClientSampleId :

PB94794BSD


Tgt Ion:128 Resp: 37023

Ion Ratio Lower Upper

128 100

129 10.8 11.2 16.8#




127 13.4 11.6 17.4



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#10

Hexachlorobutadiene

Concen: 3.80 ng

RT: 11.26 min Scan# 724

Delta R.T. -0.00 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57


Tgt Ion:225 Resp: 5448

Ion Ratio Lower Upper

225 100

223 63.1 50.6 76.0



227 63.8 51.4 77.0




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E





#11

2-Methylnaphthalene

Concen: 4.10 ng

RT: 12.60 min Scan# 812

Delta R.T. -0.04 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57

Instrument :

BNA_E

ClientSampleId :

PB94794BSD


Tgt Ion:142 Resp: 24680

Ion Ratio Lower Upper

142 100

141 87.0 73.7 110.5




115 35.1 34.0 51.0



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.80 min Scan# 979

Delta R.T. 0.00 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57


Tgt Ion:164 Resp: 29360

Ion Ratio Lower Upper

164 100

162 93.7 71.4 107.0



160 38.2 27.4 41.2




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#13
 2,4,6-Tribromophenol
 Concen: 10.39 ng
 RT: 16.28 min Scan# 1086
 Delta R.T. -0.02 min
 Lab File: BE092569.D
 Acq: 18 Nov 2016 16:57




Instrument :
 BNA_E
 ClientSampleId :
 PB94794BSD


Tgt Ion: 330 Resp: 6842
 Ion Ratio Lower Upper
 330 100
 332 97.8 75.4 113.0
 141 44.2 31.0 46.6



#14
 2-Fluorobiphenyl
 Concen: 4.26 ng
 RT: 13.41 min Scan# 882
 Delta R.T. -0.02 min
 Lab File: BE092569.D
 Acq: 18 Nov 2016 16:57

Tgt Ion: 172 Resp: 30495
 Ion Ratio Lower Upper
 172 100
 171 36.5 30.1 45.1
 170 26.1 21.8 32.6





#15

Acenaphthylene

Concen: 4.05 ng

RT: 14.51 min Scan# 962

Delta R.T. 0.00 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57

Instrument :

BNA_E

ClientSampleId :

PB94794BSD


Tgt Ion:152 Resp: 165471

Ion Ratio Lower Upper

152 100

151 4.8 3.9 5.9




153 12.9 10.4 15.6



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#16

Acenaphthene

Concen: 3.64 ng

RT: 14.85 min Scan# 982

Delta R.T. -0.02 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57


Tgt Ion:154 Resp: 24306

Ion Ratio Lower Upper

154 100

153 453.3 350.8 526.2



152 231.9 171.1 256.7




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#17
Fluorene
 Concen: 3.92 ng
 RT: 15.85 min Scan# 1048
 Delta R.T. -0.01 min
 Lab File: BE092569.D
 Acq: 18 Nov 2016 16:57




Instrument :
 BNA_E
 ClientSampleId :
 PB94794BSD


Tgt Ion:166 Resp: 32363
 Ion Ratio Lower Upper
 166 100
 165 98.9 78.2 117.4
 167 13.4 11.0 16.4



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.56 min Scan# 1153
 Delta R.T. -0.00 min
 Lab File: BE092569.D
 Acq: 18 Nov 2016 16:57

Tgt Ion:188 Resp: 53078
 Ion Ratio Lower Upper
 188 100
 94 2.9 3.2 4.8#
 80 2.3 2.6 3.8#









#19
Hexachlorobenzene
Concen: 4.64 ng
RT: 16.87 min Scan# 1123
Delta R.T. -0.00 min
Lab File: BE092569.D
Acq: 18 Nov 2016 16:57




Instrument :
BNA_E
ClientSampleId :
PB94794BSD


Tgt Ion: 284 Resp: 9732
Ion Ratio Lower Upper
284 100
142 0.0 29.1 43.7#
249 30.5 27.9 41.9



#20
Pentachlorophenol
Concen: 7.36 ng
RT: 17.21 min Scan# 1138
Delta R.T. -0.02 min
Lab File: BE092569.D
Acq: 18 Nov 2016 16:57

Tgt Ion: 266 Resp: 7243
Ion Ratio Lower Upper
266 100
268 56.4 62.5 93.7#
264 68.6 57.2 85.8





#21

Phenanthrene

Concen: 4.50 ng

RT: 17.58 min Scan# 1154

Delta R.T. -0.02 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57

Instrument :

BNA_E

ClientSampleId :

PB94794BSD





Tgt Ion:178 Resp: 53391

Ion Ratio Lower Upper

178 100

176 20.2 15.8 23.8

179 14.8 16.0 24.0#



#22

Anthracene

Concen: 4.59 ng

RT: 17.67 min Scan# 1158

Delta R.T. -0.02 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57




Tgt Ion:178 Resp: 51378


Ion Ratio Lower Upper

178 100

176 19.6 15.0 22.4

179 14.9 12.2 18.2





#23

Fluoranthene

Concen: 3.92 ng

RT: 19.58 min Scan# 1264

Delta R.T. -0.02 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57

Instrument :

BNA_E

ClientSampleId :

PB94794BSD


Tgt Ion:202 Resp: 217373

Ion Ratio Lower Upper

202 100

101 3.1 2.2 3.4




203 17.9 13.7 20.5



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#24

Chrysene-d12

Concen: 5.00 ng

RT: 21.72 min Scan# 1374

Delta R.T. -0.00 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57


Tgt Ion:240 Resp: 68066

Ion Ratio Lower Upper

240 100

120 1.5 2.6 4.0#



236 26.4 24.6 37.0




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#25
 Pyrene
 Concen: 4.27 ng
 RT: 19.95 min Scan# 1286
 Delta R.T. -0.00 min
 Lab File: BE092569.D
 Acq: 18 Nov 2016 16:57




Instrument :
 BNA_E
 ClientSampleId :
 PB94794BSD


Tgt Ion: 202 Resp: 230871
 Ion Ratio Lower Upper
 202 100
 200 5.3 4.2 6.4
 203 18.2 13.9 20.9



#26
 Terphenyl-d14
 Concen: 4.52 ng
 RT: 20.16 min Scan# 1298
 Delta R.T. -0.02 min
 Lab File: BE092569.D
 Acq: 18 Nov 2016 16:57

Tgt Ion: 244 Resp: 35205
 Ion Ratio Lower Upper
 244 100
 212 8.4 6.7 10.1
 122 4.7 3.6 5.4





#27

Benzo(a)anthracene

Concen: 7.78 ng

RT: 21.70 min Scan# 1373

Delta R.T. -0.00 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57

Instrument :

BNA_E

ClientSampleId :

PB94794BSD


Tgt Ion:228 Resp: 505385

Ion Ratio Lower Upper

228 100

226 22.9 23.6 35.4#




229 18.6 13.3 19.9



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#28

Chrysene

Concen: 8.76 ng

RT: 21.70 min Scan# 1373

Delta R.T. -0.07 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57


Tgt Ion:228 Resp: 506768

Ion Ratio Lower Upper

228 100

226 22.9 36.3 54.5#



229 18.6 10.6 16.0#




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#29
 Bis(2-ethylhexyl)phthalate
 Concen: 5.14 ng
 RT: 21.61 min Scan# 1369
 Delta R.T. -0.02 min
 Lab File: BE092569.D
 Acq: 18 Nov 2016 16:57




Instrument :
 BNA_E
 ClientSampleId :
 PB94794BSD


Tgt Ion:149 Resp: 22725
 Ion Ratio Lower Upper
 149 100
 167 22.7 16.0 24.0
 279 0.0 16.0 24.0#



#30
 Indeno(1,2,3-cd)pyrene
 Concen: 4.72 ng
 RT: 26.55 min Scan# 1713
 Delta R.T. -0.02 min
 Lab File: BE092569.D
 Acq: 18 Nov 2016 16:57

Tgt Ion:276 Resp: 298018
 Ion Ratio Lower Upper
 276 100
 138 27.7 20.3 30.5
 277 24.5 19.7 29.5









#31
Perylene-d12
 Concen: 5.00 ng
 RT: 24.08 min Scan# 1543
 Delta R.T. -0.01 min
 Lab File: BE092569.D
 Acq: 18 Nov 2016 16:57




Instrument :
 BNA_E
 ClientSampleId :
 PB94794BSD


Tgt Ion: 264 Resp: 59429
 Ion Ratio Lower Upper
 264 100
 260 25.1 20.1 30.1
 265 22.3 17.9 26.9



#32
Benzo(b)fluoranthene
 Concen: 4.13 ng
 RT: 23.35 min Scan# 1480
 Delta R.T. -0.01 min
 Lab File: BE092569.D
 Acq: 18 Nov 2016 16:57

Tgt Ion: 252 Resp: 59507
 Ion Ratio Lower Upper
 252 100
 253 23.1 18.7 28.1
 125 10.0 10.5 15.7#





#33

Benzo(k)fluoranthene

Concen: 3.75 ng

RT: 23.39 min Scan# 1484

Delta R.T. -0.01 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57

Instrument :

BNA_E

ClientSampleId :

PB94794BSD


Tgt Ion:252 Resp: 56610

Ion Ratio Lower Upper

252 100

253 22.2 20.3 30.5




125 10.1 9.6 14.4



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#34

Benzo(a)pyrene

Concen: 4.28 ng

RT: 23.96 min Scan# 1533

Delta R.T. -0.01 min

Lab File: BE092569.D

Acq: 18 Nov 2016 16:57


Tgt Ion:252 Resp: 57112

Ion Ratio Lower Upper

252 100

253 21.6 19.0 28.6



125 11.8 11.8 17.8#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#35
Dibenzo(a,h)anthracene
Concen: 4.65 ng
RT: 26.56 min Scan# 1714
Delta R.T. -0.03 min
Lab File: BE092569.D
Acq: 18 Nov 2016 16:57

Instrument :
BNA_E
ClientSampleId :
PB94794BSD

Tgt Ion: 278 Resp: 61326
Ion Ratio Lower Upper
278 100
139 17.2 13.8 20.8
279 23.6 21.4 32.2



#36
Benzo(g,h,i)perylene
Concen: 4.62 ng
RT: 27.31 min Scan# 1758
Delta R.T. -0.02 min
Lab File: BE092569.D
Acq: 18 Nov 2016 16:57

Tgt Ion: 276 Resp: 258350
Ion Ratio Lower Upper
276 100
277 23.9 19.8 29.8
138 22.3 19.8 29.6

