

Data Path : Z:\HPCHEM1\BNA_E\Data\BE111816\
 Data File : BE092563.D
 Acq On : 18 Nov 2016 12:07
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 21 04:55:18 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE111616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 15:43:18 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.681	0.769	-12.9	126	0.00
3	n-Nitrosodimethylamine	0.654	0.633	3.2	122	0.00
4 S	2-Fluorophenol	0.844	0.861	-2.0	110	0.00
5 S	Phenol-d6	1.104	1.187	-7.5	126	0.00
6	bis(2-Chloroethyl)ether	1.341	1.264	5.7	114	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.269	0.270	-0.4	113	0.00
9	Naphthalene	1.032	1.023	0.9	110	-0.02
10	Hexachlorobutadiene	0.157	0.158	-0.6	110	0.00
11	2-Methylnaphthalene	0.658	0.649	1.4	115	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
13 S	2,4,6-Tribromophenol	0.112	0.115	-2.7	127	0.00
14 S	2-Fluorobiphenyl	1.219	1.203	1.3	114	0.00
15	Acenaphthylene	6.962	6.803	2.3	118	0.00
16	Acenaphthene	1.136	1.132	0.4	116	0.00
17	Fluorene	1.408	1.409	-0.1	117	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
19	Hexachlorobenzene	0.197	0.178	9.6	116	0.00
20	Pentachlorophenol	0.036	0.034	5.6	139	0.00
21	Phenanthrene	1.119	1.072	4.2	115	0.00
22	Anthracene	1.053	1.015	3.6	119	0.00
23	Fluoranthene	5.219	5.005	4.1	118	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
25	Pyrene	3.975	4.056	-2.0	117	0.00
26 S	Terphenyl-d14	0.572	0.580	-1.4	118	0.00
27	Benzo(a)anthracene	4.692	4.648	0.9	126	0.00
28	Chrysene	4.250	4.862	-14.4	118	-0.02
29	Bis(2-ethylhexyl)phthalate	0.305	0.296	3.0	124	-0.02
30	Indeno(1,2,3-cd)pyrene	4.722	4.588	2.8	110	0.00
31 I	Perylene-d12	1.000	1.000	0.0	108	-0.01
32	Benzo(b)fluoranthene	1.213	1.279	-5.4	119	-0.01
33	Benzo(k)fluoranthene	1.270	1.133	10.8	102	0.00
34 C	Benzo(a)pyrene	1.122	1.080	3.7	113	0.00
35	Dibenzo(a,h)anthracene	1.111	1.048	5.7	110	-0.02
36	Benzo(g,h,i)perylene	4.702	4.464	5.1	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0