

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111918\
 Data File : BE098244.D
 Acq On : 19 Nov 2018 9:13
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 19 10:22:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 16 19:05:33 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.628	0.637	-1.4	110	0.00
3	n-Nitrosodimethylamine	0.706	0.746	-5.7	116	0.00
4 S	2-Fluorophenol	1.073	1.146	-6.8	109	0.00
5 S	Phenol-d6	1.685	1.811	-7.5	108	0.00
6	bis(2-Chloroethyl)ether	1.443	1.457	-1.0	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.378	0.388	-2.6	101	0.01
9	Naphthalene	4.073	4.214	-3.5	101	0.00
10	Hexachlorobutadiene	0.239	0.239	0.0	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.702	0.706	-0.6	100	0.00
12	2-Methylnaphthalene	0.718	0.723	-0.7	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.01
14 S	2,4,6-Tribromophenol	0.129	0.138	-7.0	109	0.00
15 S	2-Fluorobiphenyl	1.739	1.723	0.9	99	0.01
16	Acenaphthylene	1.904	1.974	-3.7	100	0.00
17	Acenaphthene	1.137	1.199	-5.5	100	0.00
18	Fluorene	1.449	1.531	-5.7	102	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.248	0.251	-1.2	102	0.01
21	Hexachlorobenzene	0.256	0.261	-2.0	105	0.00
22	Pentachlorophenol	0.035	0.045	-28.6#	134	0.00
23	Phenanthrene	1.029	1.077	-4.7	106	0.00
24	Anthracene	0.935	0.949	-1.5	103	0.01
25 SURR	Fluoranthene-d10	4.866	5.182	-6.5	108	0.00
26	Fluoranthene	1.235	1.298	-5.1	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
28	Pvrene	1.169	1.184	-1.3	107	0.00
29 S	Terphenyl-d14	0.919	0.927	-0.9	108	0.00
30	Benzo(a)anthracene	1.157	1.179	-1.9	111	0.00
31	Chrysene	1.191	1.225	-2.9	110	0.00
32	Bis(2-ethylhexyl)phthalate	1.804	2.116	-17.3	133	0.00
33	Indeno(1,2,3-cd)pyrene	1.019	1.090	-7.0	115	0.00
34 I	Perylene-d12	1.000	1.000	0.0	110	0.00
35	Benzo(b)fluoranthene	1.130	1.161	-2.7	114	0.00
36	Benzo(k)fluoranthene	1.276	1.293	-1.3	109	0.00
37 C	Benzo(a)pyrene	1.109	1.141	-2.9	113	0.00
38	Dibenzo(a,h)anthracene	0.889	0.912	-2.6	115	0.00
39	Benzo(a,h,i)perylene	0.958	1.014	-5.8	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE111918\
Data File : BE098244.D
Acq On : 19 Nov 2018 9:13
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 19 10:22:24 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE111618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 16 19:05:33 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			