

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112015\
 Data File : BE091255.D
 Acq On : 20 Nov 2015 20:56
 Operator : UM/NP
 Sample : G4310-12
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 D9KM7

Manual Integrations
 APPROVED

Mmdadoda
 11/23/2015 2:12:49 PM

Quant Time: Nov 23 09:37:54 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 21 02:52:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	2722	0.40	ng/ul	-0.03
4) Naphthalene-d8	9.55	136	13485	0.40	ng/ul	-0.02
8) Acenaphthene-d10	13.36	164	7549	0.40	ng/ul	-0.01
12) Phenanthrene-d10	16.08	188	18508m	0.40	ng/ul	-0.01
18) Chrysene-d12	20.23	240	21437	0.40	ng/ul	0.00
22) Perylene-d12	22.11	264	20803m	0.40	ng/ul	0.00
System Monitoring Compounds						
2) 1,4-Dioxane-d8	2.93	96	22900	11.70	ng/uL	-0.04
6) 2-Methylnaphthalene-d10	11.06	152	6511	0.35	ng/ul	0.00
16) Fluoranthene-d10	18.07	212	21500	0.41	ng/ul	0.00
Target Compounds						
13) Pentachlorophenol	15.85	266	247	0.14	ng/ul	Ovalue 97
14) Phenanthrene	16.11	178	10608m	0.05	ng/ul	
17) Fluoranthene	18.10	202	9294	0.04	ng/ul	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE112015\
Data File : BE091255.D
Acq On : 20 Nov 2015 20:56
Operator : UM/NP
Sample : G4310-12
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
D9KM7

Manual Integrations
APPROVED

Mmdadoda
11/23/2015 2:12:49 PM

Quant Time: Nov 23 09:37:54 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 21 02:52:22 2015
Response via : Initial Calibration

