

Data Path : Z:\HPCHEM1\BNA E\Data\BE112015\
 Data File : BE091246.D
 Acq On : 20 Nov 2015 13:11
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.476

Quant Time: Nov 21 02:31:03 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE102715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 19 04:25:56 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
2 SURR	1,4-Dioxane-d8	0.288	0.307	-6.6	73	0.00
3	1,4-Dioxane	0.325	0.350	-7.7	77	0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	70	0.02
5	Naphthalene	0.986	1.025	-4.0	71	0.02
6 SURR	2-Methylnaphthalene-d10	0.546	0.527	3.5	67	0.02
7	2-Methylnaphthalene	0.643	0.592	7.9	64	0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	61	0.00
9	Acenaphthylene	1.998	2.059	-3.1	61	0.00
10 C	Acenaphthene	1.368	1.467	-7.2	64	0.00
11	Fluorene	1.757	1.839	-4.7	60	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	58	0.01
13	Pentachlorophenol	0.038	0.051	-34.2#	96	0.01
14	Phenanthrene	4.619	4.454	3.6	55	0.00
15	Anthracene	4.200	4.077	2.9	58	0.00
16 SURR	Fluoranthene-d10	1.127	1.044	7.4	57	0.00
17 C	Fluoranthene	5.406	5.547	-2.6	62	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	62	0.00
19	Pyrene	4.544	4.477	1.5	60	0.00
20	Benzo(a)anthracene	1.027	0.958	6.7	59	0.00
21	Chrysene	1.183	1.240	-4.8	64	0.00
22 I	Perylene-d12	1.000	1.000	0.0	55	0.00
23	Benzo(b)fluoranthene	1.249	1.263	-1.1	55	0.00
24	Benzo(k)fluoranthene	1.281	1.535	-19.8	63	0.00
25 C	Benzo(a)pyrene	1.150	1.250	-8.7	59	0.00
26	Indeno(1,2,3-cd)pyrene	4.847	4.963	-2.4	55	0.00
27	Dibenzo(a,h)anthracene	1.155	1.170	-1.3	55	0.00
28	Benzo(q,h,i)perylene	5.083	5.182	-1.9	56	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0